

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122619\
 Data File : PS008157.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Dec 2019 18:44
 Operator : SM\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS122619

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 02:51:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122619.M
 Quant Title : 8080.M
 QLast Update : Fri Dec 27 02:43:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	2.401	2.357 E6	1.8	99	0.00
2 T	3,5-DICHLOROBENZOIC ACID	2.484	2.517 E6	-1.3	99	0.00
3 T	4-Nitrophenol	683.394	745.278 E3	-9.1	103	0.02
4 S	2,4-DCAA	2.020	1.996 E6	1.2	99	0.00
5 T	DICAMBA	7.272	7.262 E6	0.1	99	0.00
6 T	MCPD	5.454	5.515 E6	-1.1	99	0.00
7 T	MCPA	7.648	7.679 E6	-0.4	98	0.00
8 T	DICHLOROPROP	1.772	1.761 E6	0.6	98	0.00
9 T	2,4-D	1.641	1.631 E6	0.6	97	0.00
10 T	Pentachlorophenol	26.191	25.981 E6	0.8	98	0.00
11 T	2,4,5-TP (SILVEX)	10.315	10.253 E6	0.6	98	0.00
12 T	2,4,5-T	12.016	11.618 E6	3.3	99	0.00
13 T	2,4-DB	1.329	1.320 E6	0.7	98	0.00
14 T	DINOSEB	9.690	9.341 E6	3.6	98	0.00
15 T	Picloram	10.861	11.373 E6	-4.7	98	0.01
16 T	DCPA	13.569	13.302 E6	2.0	100	0.00

Signal #2

1 T	Dalapon	1.020	1.000 E6	2.0	99	0.00
2 T	3,5-DICHLOROBENZOIC ACID	1.164	1.157 E6	0.6	99	0.00
3 T	4-Nitrophenol	495.357	497.006 E3	-0.3	102	0.00
4 S	2,4-DCAA	895.382	871.343 E3	2.7	98	0.00
5 T	DICAMBA	3.407	3.339 E6	2.0	98	0.00
6 T	MCPD	2.588	2.563 E6	1.0	99	0.00
7 T	MCPA	3.804	3.699 E6	2.8	99	0.00
8 T	DICHLOROPROP	856.026	834.974 E3	2.5	98	0.00
9 T	2,4-D	1.015	0.987 E6	2.8	99	0.00
10 T	Pentachlorophenol	12.726	12.374 E6	2.8	98	0.00
11 T	2,4,5-TP (SILVEX)	4.997	4.880 E6	2.3	98	0.00
12 T	2,4,5-T	4.180	4.130 E6	1.2	97	0.00
13 T	2,4-DB	625.576	638.982 E3	-2.1	103	0.00
14 T	DINOSEB	3.745	3.660 E6	2.3	98	0.00
15 T	Picloram	6.527	6.678 E6	-2.3	98	0.00
16 T	DCPA	6.671	6.513 E6	2.4	99	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122619\
Data File : PS008157.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2019 18:44
Operator : SM\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS122619

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 02:51:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122619.M
Quant Title : 8080.M
QLast Update : Fri Dec 27 02:43:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0