

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122619\
 Data File : PS008152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Dec 2019 16:54
 Operator : SM\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

Ankita
 12/27/2019 8:32:51 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 02:22:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122619.M
 Quant Title : 8080.M
 QLast Update : Fri Dec 27 02:18:01 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.356	5.093	459.4E6	184.0E6	228.352	207.889m
Target Compounds							
1) T	Dalapon	1.941	1.661	445.1E6	180.0E6	187.165m	177.864m
2) T	3,5-DICHL...	4.743	4.407	479.6E6	218.8E6	189.124	186.458
3) T	4-Nitroph...	5.256	4.799	92119184	83306099	127.582	171.663 #
5) T	DICAMBA	5.489	5.211	1525.4E6	674.5E6	207.575	198.964
6) T	MCPD	5.628	5.288	96918558	46070720	17.337	17.821
7) T	MCPA	5.755	5.459	139.3E6	73619267	17.853m	19.752
8) T	DICHLORPROP	6.112	5.735	363.8E6	165.1E6	203.378	194.635
9) T	2,4-D	6.357	6.010	308.6E6	199.8E6	183.906m	199.649
10) T	Pentachlo...	6.567	6.352	5888.1E6	2810.4E6	222.787m	223.418
11) T	2,4,5-TP ...	7.109	6.718	2199.7E6	1018.1E6	210.668	204.968
12) T	2,4,5-T	7.399	7.079	2627.3E6	747.1E6	223.400	175.841
13) T	2,4-DB	7.929	7.562	225.7E6	113.9E6	167.987	183.807
14) T	DINOSEB	8.970	7.858	2226.3E6	744.6E6	234.514	200.169
15) T	Picloram	8.864	8.802	1264.0E6	1044.6E6	108.791	153.865 #
16) T	DCPA	9.245	8.744	2932.9E6	1404.6E6	220.007	214.165

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122619\
Data File : PS008152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2019 16:54
Operator : SM\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Manual Integrations
APPROVED

Ankita
12/27/2019 8:32:51 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 02:22:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122619.M
Quant Title : 8080.M
QLast Update : Fri Dec 27 02:18:01 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

