

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122722\
 Data File : PS021563.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2022 15:53
 Operator : AR\AJ
 Sample : N6051-02MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SP-591MSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 12/28/2022
 Supervised By : Ankita Jodhani 12/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 21:09:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 07 15:59:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.373	6.807	1238.2E6	920.4E6	426.159m	533.870 #
Target Compounds							
1) T	Dalapon	2.480	2.555	885.0E6	1195.4E6	280.051m	532.494 #
2) T	3,5-DICHL...	5.746	5.989	1971.9E6	1289.5E6	496.149	492.797
3) T	4-Nitroph...	6.224	6.432	21725944	16706476	12.083m	14.336
5) T	DICAMBA	6.516	6.972	5112.8E6	3540.6E6	444.306	463.600
6) T	MCP	6.671	7.069	392.7E6	302.1E6	42.057	51.351
7) T	MCPA	6.795	7.280	565.4E6	478.7E6	44.139m	54.134m
8) T	DICHLORPROP	7.113	7.596	1463.9E6	1087.2E6	465.640	551.988
9) T	2,4-D	7.314	7.887	1786.2E6	1262.3E6	467.453	523.289
10) T	Pentachlo...	7.543	8.315	21198.2E6	13588.7E6	521.279	515.461
11) T	2,4,5-TP ...	8.072	8.677	9050.7E6	6044.7E6	489.475	539.776
12) T	2,4,5-T	8.331	9.053	8937.1E6	5289.7E6	473.676	488.047
13) T	2,4-DB	8.847	9.561	1931.3E6	927.7E6	527.357	645.373
14) T	DINOSEB	9.894	9.886	6369.1E6	3448.2E6	399.749	435.976
15) T	Picloram	9.729	10.857	11381.6E6	6782.8E6	397.401	417.105
16) T	DCPA	10.161	10.814	11772.9E6	6429.4E6	462.268	473.858

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122722\
Data File : PS021563.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2022 15:53
Operator : AR\AJ
Sample : N6051-02MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SP-591MSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/28/2022
Supervised By : Ankita Jodhani 12/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 21:09:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
Quant Title : 8080.M
QLast Update : Wed Dec 07 15:59:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

