

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122822\
 Data File : PS021591.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Dec 2022 23:36
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 06:39:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 07 15:59:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.374	6.807	2081.0E6	1350.3E6	716.237	783.267
Target Compounds							
1) T	Dalapon	2.478	2.555	2145.3E6	1580.0E6	678.880	703.817
2) T	3,5-DICHL...	5.746	5.989	2792.4E6	1942.4E6	702.606	742.281
3) T	4-Nitroph...	6.213	6.423	1170.2E6	822.9E6	650.784	706.165
5) T	DICAMBA	6.515	6.971	8429.1E6	5779.6E6	732.503	756.773
6) T	MCP	6.673	7.072	634.7E6	448.1E6	67.981	76.153
7) T	MCPA	6.798	7.284	911.7E6	622.4E6	71.174	70.387
8) T	DICHLORPROP	7.112	7.596	2213.1E6	1447.0E6	703.959	734.673
9) T	2,4-D	7.315	7.888	2654.7E6	1768.9E6	694.740	733.282
10) T	Pentachlo...	7.543	8.315	30735.5E6	19851.6E6	755.808	753.033
11) T	2,4,5-TP ...	8.071	8.677	13333.7E6	8457.0E6	721.109	755.187
12) T	2,4,5-T	8.331	9.053	13657.0E6	8180.8E6	723.838	754.787
13) T	2,4-DB	8.839	9.559	2458.6E6	1104.8E6	671.333	768.549
14) T	DINOSEB	9.893	9.887	11401.3E6	5977.2E6	715.592	755.733
15) T	Picloram	9.728	10.858	17711.2E6	10393.5E6	618.408	639.144
16) T	DCPA	10.160	10.815	18425.2E6	10226.0E6	723.477	753.668

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122822\
Data File : PS021591.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Dec 2022 23:36
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 06:39:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
Quant Title : 8080.M
QLast Update : Wed Dec 07 15:59:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

