

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123021\
 Data File : PS017864.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Dec 2021 09:50
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 00:09:48 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
 Quant Title : 8080.M
 QLast Update : Tue Dec 14 13:29:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.125	6.713	1199.5E6	478.0E6	767.474	801.843
Target Compounds							
1) T	Dalapon	2.976	2.482	1755.2E6	559.6E6	684.911	722.864
2) T	3,5-DICHL...	6.353	5.917	1642.6E6	703.0E6	735.207	775.363
3) T	4-Nitroph...	6.959	6.347	428.5E6	249.8E6	633.718	717.318
5) T	DICAMBA	7.312	6.870	4828.1E6	2099.2E6	763.528	788.017
6) T	MCPD	7.482	6.966	355.7E6	170.7E6	75.450	78.434
7) T	MCPA	7.629	7.177	497.6E6	233.4E6	75.238	76.335
8) T	DICHLORPROP	8.002	7.487	1190.0E6	527.1E6	761.629	773.438
9) T	2,4-D	8.233	7.781	1241.2E6	627.1E6	702.107	771.347
10) T	Pentachlo...	8.533	8.196	16492.2E6	6745.5E6	764.330	787.307
11) T	2,4,5-TP ...	9.082	8.558	6530.7E6	2916.4E6	767.698	802.237
12) T	2,4,5-T	9.367	8.935	6522.0E6	2626.5E6	749.720	768.059
13) T	2,4-DB	9.917	9.439	989.7E6	338.6E6	670.186	747.300
14) T	DINOSEB	11.061	9.758	5131.5E6	2022.0E6	750.453	782.645
15) T	Picloram	10.889	10.743	7316.1E6	3389.1E6	674.218	690.041
16) T	DCPA	11.329	10.685	8383.9E6	3618.6E6	771.966	807.930

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123021\
Data File : PS017864.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Dec 2021 09:50
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 00:09:48 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
Quant Title : 8080.M
QLast Update : Tue Dec 14 13:29:21 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

