

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123119\
 Data File : PS008256.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Dec 2019 11:12
 Operator : SM\AJ
 Sample : PB125807BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB125807BS

Manual Integrations
 APPROVED

mohammad
 1/2/2020 12:37:59 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 11:39:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122619.M
 Quant Title : 8080.M
 QLast Update : Fri Dec 27 02:43:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.367	5.132	971.4E6	412.1E6	480.800	460.302
Target Compounds							
1) T	Dalapon	1.968	1.706	1006.7E6	439.8E6	419.369	431.090
2) T	3,5-DICHL...	4.752	4.444	1097.2E6	506.3E6	441.651	434.813
3) T	4-Nitroph...	5.235	4.828	320.1E6	208.1E6	468.326	420.104
5) T	DICAMBA	5.509	5.253	3281.9E6	1503.9E6	451.338	441.415
6) T	MCPD	5.656	5.335	227.5E6	108.4E6	41.713	41.891
7) T	MCPA	5.784	5.506	325.9E6	158.3E6	42.617	41.621
8) T	DICHLORPROP	6.125	5.773	791.4E6	374.0E6	446.665	436.884
9) T	2,4-D	6.342	6.041	798.5E6	434.2E6	486.583	427.642
10) T	Pentachlo...	6.587	6.394	12140.8E6	5751.2E6	463.551	451.933
11) T	2,4,5-TP ...	7.121	6.757	4702.5E6	2206.6E6	455.904	441.578
12) T	2,4,5-T	7.394	7.113	5381.5E6	1989.4E6	447.847	475.963
13) T	2,4-DB	7.920	7.595	611.8E6	277.5E6	460.438	443.651
14) T	DINOSEB	8.992	7.898	3914.0E6	1596.3E6	403.903	426.225
15) T	Picloram	8.834	8.831	5914.6E6	2942.5E6	535.402	450.846m
16) T	DCPA	9.266	8.787	6026.7E6	2992.2E6	444.155	448.545

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123119\
Data File : PS008256.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Dec 2019 11:12
Operator : SM\AJ
Sample : PB125807BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB125807BS

Manual Integrations
APPROVED

mohammad
1/2/2020 12:37:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 11:39:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122619.M
Quant Title : 8080.M
QLast Update : Fri Dec 27 02:43:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

