

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123120\
 Data File : PS013636.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Dec 2020 13:50
 Operator : DD\AJ
 Sample : L5242-03
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 14:50:14 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 03 02:59:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.208	6.911	248.5E6	116.2E6	266.898	295.470
Target Compounds							
1) T	Dalapon	2.990f	2.587	72574720	45500134	67.561	86.579 #
2) T	3,5-DICHL...	0.000	6.094	0	21993102	N.D.	35.395
5) T	DICAMBA	0.000	7.048	0	5494674	N.D.	3.269
6) T	MCP	7.561	0.000	9135287	0	3.364	N.D. #
7) T	MCPA	7.753	7.371	47433647	32916233	12.136	15.828 #
8) T	DICHLORPROP	8.102	7.705	10832861	5178101	11.704	11.586
9) T	2,4-D	8.327	0.000	23189818	0	23.036	N.D. #
10) T	Pentachlo...	8.693f	8.456	86389502	64547630	6.446	11.447 #
11) T	2,4,5-TP ...	9.165	8.757	18868077	14696823	3.539	6.260 #
12) T	2,4,5-T	9.510f	9.158	40105596	4557374	7.254	2.043 #
13) T	2,4-DB	10.068f	9.706	142.1E6	71343682	154.705	243.920 #
14) T	DINOSEB	11.172	10.017	71989885	18937462	16.987	11.457 #
15) T	Picloram	10.962f	11.013	113.5E6	41604352	16.474	11.949 #
16) T	DCPA	11.498f	10.925	5794125	14508935	<MDL	4.771 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123120\
Data File : PS013636.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Dec 2020 13:50
Operator : DD\AJ
Sample : L5242-03
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 14:50:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
Quant Title : 8080.M
QLast Update : Thu Dec 03 02:59:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.990 min
Delta R.T.: -0.041 min
Response: 72574720
Conc: 67.56 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.587 min
Delta R.T.: -0.035 min
Response: 45500134
Conc: 86.58 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.428 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.094 min
Delta R.T.: 0.026 min
Response: 21993102
Conc: 35.39 ng/ml



#4 2,4-DCAA


R.T.: 7.208 min
Delta R.T.: -0.003 min
Response: 248498072
Conc: 266.90 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.911 min
Delta R.T.: -0.003 min
Response: 116226668
Conc: 295.47 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 7.401 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 7.048 min
Delta R.T.: -0.032 min
Response: 5494674
Conc: 3.27 ng/ml



#6 MCPP


R.T.: 7.561 min
Delta R.T.: -0.006 min
Response: 9135287
Conc: 3.36 ug/ml

Instrument :
ECD_S
ClientSampleId :




#6 MCPP

R.T.: 0.000 min
Exp R.T. : 7.174 min
Response: 0
Conc: N.D.




#7 MCPA

R.T.: 7.753 min
Delta R.T.: 0.036 min
Response: 47433647
Conc: 12.14 ug/ml



#7 MCPA


R.T.: 7.371 min
Delta R.T.: -0.018 min
Response: 32916233
Conc: 15.83 ug/ml



#8 DICHLORPROP


R.T.: 8.102 min
Delta R.T.: 0.011 min
Response: 10832861
Conc: 11.70 ng/ml

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 7.705 min
Delta R.T.: 0.000 min
Response: 5178101
Conc: 11.59 ng/ml




#9 2,4-D

R.T.: 8.327 min
Delta R.T.: 0.003 min
Response: 23189818
Conc: 23.04 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 8.004 min
Response: 0
Conc: N.D.



#10 Pentachlorophenol


R.T.: 8.693 min
Delta R.T.: 0.062 min
Response: 86389502
Conc: 6.45 ng/ml

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 8.456 min
Delta R.T.: 0.029 min
Response: 64547630
Conc: 11.45 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.165 min
Delta R.T.: -0.014 min
Response: 18868077
Conc: 3.54 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 8.757 min
Delta R.T.: -0.032 min
Response: 14696823
Conc: 6.26 ng/ml



#12 2,4,5-T


R.T.: 9.510 min
Delta R.T.: 0.043 min
Response: 40105596
Conc: 7.25 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 9.158 min
Delta R.T.: -0.011 min
Response: 4557374
Conc: 2.04 ng/ml




#13 2,4-DB

R.T.: 10.068 min
Delta R.T.: 0.047 min
Response: 142116008
Conc: 154.71 ng/ml



#13 2,4-DB


R.T.: 9.706 min
Delta R.T.: 0.028 min
Response: 71343682
Conc: 243.92 ng/ml



#14 DINOSEB


R.T.: 11.172 min
Delta R.T.: 0.000 min
Response: 71989885
Conc: 16.99 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 10.017 min
Delta R.T.: 0.013 min
Response: 18937462
Conc: 11.46 ng/ml




#15 Picloram

R.T.: 10.962 min
Delta R.T.: -0.047 min
Response: 113471216
Conc: 16.47 ng/ml



#15 Picloram


R.T.: 11.013 min
Delta R.T.: 0.012 min
Response: 41604352
Conc: 11.95 ng/ml



#16 DCPA

R.T.: 11.498 min
Delta R.T.: 0.061 min
Response: 5794125
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 10.925 min
Delta R.T.: -0.005 min
Response: 14508935
Conc: 4.77 ng/ml