

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123120\
 Data File : PS013644.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Dec 2020 16:54
 Operator : DD\AJ
 Sample : BS TEST
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BS TEST

Manual Integrations
 APPROVED

Ankita
 1/4/2021 2:26:59 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 18:05:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 03 02:59:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.207	6.909	381.1E6	170.0E6	409.324	432.281
Target Compounds							
1) T	Dalapon	3.028	2.622	464.9E6	235.0E6	432.791	447.079
2) T	3,5-DICHL...	6.424	6.064	525.0E6	237.1E6	409.647	381.618
3) T	4-Nitroph...	7.041	6.522	179.9E6	93216809	399.249	391.093
5) T	DICAMBA	7.398	7.076	1442.4E6	691.5E6	390.602	411.440
6) T	MCPD	7.563	7.169	118.5E6	55007667	43.656	41.510
7) T	MCPA	7.711	7.382	171.7E6	103.1E6	43.938	49.560m
8) T	DICHLORPROP	8.088	7.701	389.9E6	182.1E6	421.298	407.416
9) T	2,4-D	8.320	7.997	512.2E6	240.3E6	508.852	445.326
10) T	Pentachlo...	8.626	8.422	5111.8E6	2245.2E6	381.407	398.157
11) T	2,4,5-TP ...	9.174	8.783	2257.5E6	1005.2E6	423.481	428.159
12) T	2,4,5-T	9.460	9.164	2179.7E6	959.8E6	394.239	430.176
13) T	2,4-DB	10.013	9.669	251.8E6	128.9E6	274.137	440.831 #
14) T	DINOSEB	11.168	10.000	258.2E6	92862691	60.920	56.182
15) T	Picloram	10.994	10.986	2664.1E6	1397.5E6	386.781m	401.380
16) T	DCPA	11.435	10.925	2454.7E6	1363.7E6	361.961	448.416

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123120\
Data File : PS013644.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Dec 2020 16:54
Operator : DD\AJ
Sample : BS TEST
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
BS TEST

Manual Integrations
APPROVED

Ankita
1/4/2021 2:26:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 18:05:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
Quant Title : 8080.M
QLast Update : Thu Dec 03 02:59:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

