

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123124\
 Data File : PS028862.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Dec 2024 08:59
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 01/02/2025
 Supervised By :Ankita Jodhani 01/02/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 01 00:24:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122324.M
 Quant Title : 8080.M
 QLast Update : Mon Dec 23 13:44:25 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.207	7.690	1787.7E6	778.4E6	790.328m	691.612
Target Compounds							
1) T	Dalapon	2.620	2.672	1939.8E6	1304.7E6	688.705	662.780
2) T	3,5-DICHL...	6.383	6.653	2439.0E6	1071.1E6	740.194	646.181
3) T	4-Nitroph...	7.006	7.219	1060.8E6	530.3E6	706.227	633.597
5) T	DICAMBA	7.393	7.887	7469.3E6	3669.1E6	757.487	668.348
6) T	MCPD	7.575	7.991	460.9E6	214.4E6	74.007	65.624
7) T	MCPA	7.724	8.233	627.0E6	297.7E6	71.533	64.253
8) T	DICHLORPROP	8.098	8.600	1973.8E6	916.2E6	743.937	664.047
9) T	2,4-D	8.327	8.928	2114.4E6	953.7E6	742.006m	660.874m
10) T	Pentachlo...	8.624	9.452	30444.1E6	14903.4E6	779.278	683.683
11) T	2,4,5-TP ...	9.201	9.828	12021.4E6	6090.8E6	765.243	677.996
12) T	2,4,5-T	9.493	10.246	12262.5E6	5794.2E6	753.725	665.298
13) T	2,4-DB	10.063	10.811	2237.7E6	639.7E6	716.851	658.688
14) T	DINOSEB	11.268	11.189	9950.1E6	3955.8E6	740.241	642.256
15) T	Picloram	11.079	12.274	19443.5E6	8346.8E6	725.159	650.113
16) T	DCPA	11.563	12.228	18209.2E6	7336.4E6	757.108	682.802m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS123124\
Data File : PS028862.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Dec 2024 08:59
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/02/2025
Supervised By : Ankita Jodhani 01/02/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 01 00:24:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122324.M
Quant Title : 8080.M
QLast Update : Mon Dec 23 13:44:25 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

