

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012022\
 Data File : PS018087.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jan 2022 12:52
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc : FOR ABDUL RAUF
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 00:21:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 06 10:21:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.047	6.646	1028.0E6	410.7E6	711.188	707.801
Target Compounds							
1) T	Dalapon	2.908	2.419	1496.6E6	458.1E6	610.072	599.026
2) T	3,5-DICHL...	6.287	5.862	1382.7E6	586.4E6	642.727	662.840
3) T	4-Nitroph...	6.875	6.282	583.4E6	224.4E6	722.357	732.164
5) T	DICAMBA	7.233	6.800	4034.9E6	1713.7E6	646.161	651.846
6) T	MCPD	7.402	6.894	303.1E6	136.7E6	64.921	63.555
7) T	MCPA	7.550	7.104	426.6E6	191.6E6	63.945	64.200
8) T	DICHLORPROP	7.920	7.411	1004.4E6	430.7E6	638.510	645.325
9) T	2,4-D	8.147	7.701	1205.3E6	509.1E6	661.940	691.813
10) T	Pentachlo...	8.446	8.113	13863.3E6	5524.0E6	656.216	655.871
11) T	2,4,5-TP ...	8.995	8.475	5543.5E6	2364.6E6	655.098	663.338
12) T	2,4,5-T	9.278	8.851	5851.5E6	2268.3E6	660.131	693.443
13) T	2,4-DB	9.826	9.354	1170.4E6	298.1E6	708.038	673.756
14) T	DINOSEB	10.972	9.675	4192.6E6	1613.6E6	638.442	646.746
15) T	Picloram	10.798	10.654	7930.8E6	3113.2E6	664.446	665.751
16) T	DCPA	11.240	10.601	7123.5E6	2907.4E6	659.296	658.450

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012022\
Data File : PS018087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jan 2022 12:52
Operator : AJ\MA
Sample : HSTDCCC750
Misc : FOR ABDUL RAUF
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 00:21:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
Quant Title : 8080.M
QLast Update : Thu Jan 06 10:21:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

