

Data Path : P:\HPCHEM1\ECD_S\Data\PS020118\
 Data File : PS000020.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2018 21:45
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

ugo
 2/27/2018 12:06:29 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 07:03:06 2018
 Quant Method : P:\HPCHEM1\ECD_S\Method\PS020118.M
 Quant Title : 8080.M
 QLast Update : Sat Feb 03 09:57:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.860	11.239	274.9E6	234.4E6	897.612m	824.789m
Target Compounds							
1) T	Dalapon	3.200	2.576	333.0E6	264.8E6	848.054m	811.759m
2) T	3,5-DICHL...	10.272	9.380	371.3E6	303.6E6	868.745m	782.085m
3) T	4-Nitroph...	11.456	10.384	160.5E6	124.5E6	806.908m	740.528m
5) T	DICAMBA	12.180	11.529	1010.3E6	856.4E6	871.233m	777.385m
6) T	MCPD	12.520	11.755	86546583	70369292	78.591m	79.640m
7) T	MCPA	12.772	12.134	149.0E6	122.2E6	71.142m	79.237m
8) T	DICHLORPROP	13.372	12.654	244.3E6	207.2E6	866.692m	784.475
9) T	2,4-D	13.730	13.126	303.9E6	261.4E6	855.069m	810.543m
10) T	Pentachlo...	14.130	13.690	4087.0E6	3539.7E6	759.155	738.816m
11) T	2,4,5-TP ...	14.995	14.299	1401.7E6	1238.3E6	872.926m	821.301m
12) T	2,4,5-T	15.393	14.831	1302.3E6	1119.9E6	881.136	841.050m
13) T	2,4-DB	16.148	15.519	179.7E6	143.7E6	904.135m	823.047m
14) T	DINOSEB	17.661	15.952	866.9E6	733.8E6	884.250m	814.072m
15) T	Picloram	17.415	17.199	1646.5E6	1421.7E6	916.884m	844.840
16) T	DCPA	18.001	17.141	1878.4E6	1620.1E6	835.122	786.699

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_S\Data\PS020118\
Data File : PS000020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2018 21:45
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleID :
HSTDCCC750

Manual Integrations
APPROVED

ugo
2/27/2018 12:06:29 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 07:03:06 2018
Quant Method : P:\HPCHEM1\ECD_S\Method\PS020118.M
Quant Title : 8080.M
QLast Update : Sat Feb 03 09:57:21 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

