

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020322\
 Data File : PS018215.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2022 10:27
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 22:20:17 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 06 10:21:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.043	6.650	1088.4E6	437.9E6	752.981	754.607
Target Compounds							
1) T	Dalapon	2.904	2.418	1554.0E6	476.4E6	633.442	623.036
2) T	3,5-DICHL...	6.285	5.866	1462.1E6	620.3E6	679.676	701.252
3) T	4-Nitroph...	6.875	6.288	604.3E6	232.7E6	748.153	759.132
5) T	DICAMBA	7.229	6.804	4283.4E6	1824.0E6	685.965	693.773
6) T	MCPD	7.398	6.898	324.7E6	145.4E6	69.536	67.611
7) T	MCPA	7.545	7.108	455.7E6	202.2E6	68.298	67.754
8) T	DICHLORPROP	7.914	7.414	1062.5E6	455.4E6	675.454	682.329
9) T	2,4-D	8.144	7.706	1275.5E6	532.5E6	700.442	723.548
10) T	Pentachlo...	8.440	8.116	14477.6E6	5780.4E6	685.298	686.311
11) T	2,4,5-TP ...	8.990	8.480	5846.4E6	2508.0E6	690.890	703.569
12) T	2,4,5-T	9.274	8.856	6314.2E6	2389.9E6	712.333	730.629
13) T	2,4-DB	9.822	9.359	1214.5E6	326.9E6	734.680	738.723
14) T	DINOSEB	10.963	9.679	4477.8E6	1638.9E6	681.873	656.854
15) T	Picloram	10.803	10.670	7363.8E6	2955.6E6	616.947	632.061
16) T	DCPA	11.232	10.604	7388.9E6	3046.9E6	683.857	690.057

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020322\
Data File : PS018215.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2022 10:27
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 22:20:17 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
Quant Title : 8080.M
QLast Update : Thu Jan 06 10:21:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

