

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020322\
 Data File : PS018223.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2022 14:06
 Operator : AJ\MA
 Sample : N1335-10MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-4DMSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 22:24:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 06 10:21:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.039	6.647	832.7E6	351.0E6	576.076	604.974m
Target Compounds							
1) T	Dalapon	2.903	2.414	573.9E6	140.6E6	233.924	183.866
2) T	3,5-DICHL...	6.282	5.865	640.6E6	247.0E6	297.787	279.239
3) T	4-Nitroph...	6.870	6.287	249.6E6	82720779	309.053	269.869m
5) T	DICAMBA	7.226	6.803	1683.2E6	729.6E6	269.558	277.520
6) T	MCPD	7.392	6.894	137.0E6	57777533	29.335	26.862
7) T	MCPA	7.539	7.101	221.5E6	89446677	33.206	29.976
8) T	DICHLORPROP	7.911	7.413	447.2E6	193.9E6	284.259	290.510
9) T	2,4-D	8.140	7.703	583.0E6	201.8E6	320.169	274.192
10) T	Pentachlo...	8.438	8.115	6267.3E6	2546.3E6	296.664	302.331
11) T	2,4,5-TP ...	8.987	8.476	2511.4E6	1009.8E6	296.775	283.289m
12) T	2,4,5-T	9.269	8.855	2500.0E6	986.3E6	282.039	301.518
13) T	2,4-DB	9.819	9.357	367.2E6	135.6E6	222.119	306.554 #
14) T	DINOSEB	10.962	9.678	314.3E6	140.8E6	47.855	56.438
15) T	Picloram	10.794	10.664	3260.4E6	1335.9E6	273.158	285.687
16) T	DCPA	11.231	10.603	2680.7E6	1327.0E6	248.107	300.533

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020322\
Data File : PS018223.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2022 14:06
Operator : AJ\MA
Sample : N1335-10MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MH-4DMSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 22:24:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
Quant Title : 8080.M
QLast Update : Thu Jan 06 10:21:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

