

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020322\
 Data File : PS018235.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2022 20:30
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 22:29:17 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 06 10:21:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.045	6.651	1059.1E6	439.7E6	732.717	757.763
Target Compounds							
1) T	Dalapon	2.905	2.416	1566.8E6	478.0E6	638.685	625.058
2) T	3,5-DICHL...	6.286	5.867	1463.9E6	621.7E6	680.482	702.748
3) T	4-Nitroph...	6.879	6.291	580.1E6	222.4E6	718.191	725.682
5) T	DICAMBA	7.230	6.804	4325.0E6	1846.4E6	692.624	702.324
6) T	MCPD	7.398	6.898	324.8E6	147.9E6	69.560	68.758
7) T	MCPA	7.546	7.108	457.8E6	204.5E6	68.618	68.522
8) T	DICHLORPROP	7.917	7.415	1073.9E6	461.2E6	682.719	691.085
9) T	2,4-D	8.148	7.708	1278.8E6	523.3E6	702.257	711.018
10) T	Pentachlo...	8.440	8.116	14611.2E6	5870.6E6	691.618	697.027
11) T	2,4,5-TP ...	8.992	8.481	5913.1E6	2556.0E6	698.769	717.025
12) T	2,4,5-T	9.276	8.857	6461.9E6	2392.8E6	728.997	731.531
13) T	2,4-DB	9.827	9.360	1201.7E6	331.3E6	726.961	748.692
14) T	DINOSEB	10.964	9.680	4676.8E6	1682.3E6	712.168	674.266
15) T	Picloram	10.808	10.674	7319.1E6	2966.8E6	613.202	634.447
16) T	DCPA	11.234	10.604	7537.5E6	3107.1E6	697.603	703.687

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020322\
Data File : PS018235.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2022 20:30
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 22:29:17 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
Quant Title : 8080.M
QLast Update : Thu Jan 06 10:21:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

