

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021422\
 Data File : PS018324.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Feb 2022 10:05
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 15 00:27:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020822.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 09 15:57:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.028	6.638	1196.1E6	482.1E6	760.369	768.695
Target Compounds							
1) T	Dalapon	2.888	2.407	1727.1E6	529.7E6	687.460	660.977
2) T	3,5-DICHL...	6.271	5.856	1610.5E6	686.5E6	719.918	723.332
3) T	4-Nitroph...	6.861	6.278	577.5E6	285.8E6	654.495	705.598
5) T	DICAMBA	7.211	6.790	4728.3E6	2010.7E6	748.537	741.210
6) T	MCP	7.380	6.884	354.1E6	162.4E6	76.951	74.898
7) T	MCPA	7.527	7.094	493.1E6	226.6E6	75.680	73.408
8) T	DICHLORPROP	7.896	7.400	1163.9E6	507.0E6	737.648	731.496
9) T	2,4-D	8.127	7.692	1327.5E6	599.5E6	720.945	723.227
10) T	Pentachlo...	8.418	8.098	15960.1E6	6329.2E6	755.258	733.205
11) T	2,4,5-TP ...	8.968	8.463	6439.9E6	2765.6E6	753.753	746.631
12) T	2,4,5-T	9.254	8.840	6762.5E6	2620.1E6	764.170	740.000
13) T	2,4-DB	9.802	9.341	1201.2E6	361.9E6	682.482	760.136
14) T	DINOSEB	10.940	9.660	5039.2E6	1893.3E6	754.667	750.229
15) T	Picloram	10.774	10.643	7806.1E6	3517.8E6	683.254	701.882
16) T	DCPA	11.209	10.584	8272.1E6	3386.5E6	750.487	744.109

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021422\
Data File : PS018324.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Feb 2022 10:05
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 15 00:27:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020822.M
Quant Title : 8080.M
QLast Update : Wed Feb 09 15:57:26 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

