

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022221\
 Data File : PS014053.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2021 18:44
 Operator : DD\AJ
 Sample : M1422-01MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BU-03-021621MSD

Manual Integrations
 APPROVED

Ankita
 2/23/2021 9:28:23 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 02:40:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 16 01:39:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.204	6.901	417.3E6	177.9E6	416.213	465.053m
Target Compounds							
1) T	Dalapon	3.025	2.618	615.1E6	312.2E6	527.764	608.382m
2) T	3,5-DICHL...	6.421	6.058	547.3E6	220.3E6	397.921	381.005
3) T	4-Nitroph...	7.037	6.515	245.6E6	109.9E6	481.740	448.530
5) T	DICAMBA	7.393	7.067	1720.9E6	718.2E6	429.115	440.626
6) T	MCPD	7.558	7.158	113.8E6	46708374	39.032	35.722
7) T	MCPA	7.707	7.372	165.0E6	100.8E6	39.150m	52.228 #
8) T	DICHLORPROP	8.083	7.691	377.5E6	164.3E6	373.168	389.841
9) T	2,4-D	8.315	7.988	517.6E6	250.9E6	442.727m	478.383m
10) T	Pentachlo...	8.622	8.413	4498.6E6	1752.3E6	306.907	320.504
11) T	2,4,5-TP ...	9.170	8.773	2022.7E6	834.6E6	347.054	369.951
12) T	2,4,5-T	9.456	9.154	2262.1E6	862.7E6	378.560	395.374
13) T	2,4-DB	10.008	9.660	229.6E6	98341435	214.211m	334.287 #
14) T	DINOSEB	11.163	9.989	944.8E6	324.7E6	210.101	212.188
15) T	Picloram	10.992	10.979	3850.1E6	1622.4E6	467.668	495.519
16) T	DCPA	11.429	10.915	3546.3E6	1594.3E6	476.351	547.075

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS022221\
Data File : PS014053.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2021 18:44
Operator : DD\AJ
Sample : M1422-01MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
BU-03-021621MSD

Manual Integrations
APPROVED

Ankita
2/23/2021 9:28:23 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 23 02:40:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021521.M
Quant Title : 8080.M
QLast Update : Tue Feb 16 01:39:40 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

