

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS033125\
 Data File : PS029628.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2025 03:24
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDICC200

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 04/01/2025

Supervised By :mohammad ahmed 04/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 05:09:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS033125.M
 Quant Title : 8080.M
 QLast Update : Tue Apr 01 05:08:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.008	7.537	529.9E6	306.1E6	206.433	222.535
Target Compounds							
1) T	Dalapon	2.496	2.568	918.6E6	697.0E6	155.560	210.122 #
2) T	3,5-DICHL...	6.204	6.526	832.8E6	415.6E6	190.913	208.220
3) T	4-Nitroph...	6.802	7.073	206.7E6	292.3E6	192.458	205.187
5) T	DICAMBA	7.185	7.724	2658.4E6	1520.7E6	185.664	189.229
6) T	MCPD	7.504	7.828	116.0E6	74291103	14.106	18.785 #
7) T	MCPA	7.504	8.060	116.0E6	94446451	13.956	18.878 #
8) T	DICHLORPROP	7.871	8.424	544.1E6	399.9E6	196.453	211.792
9) T	2,4-D	8.096	8.742	483.4E6	437.2E6	207.428	211.714
10) T	Pentachlo...	8.372	9.245	10234.2E6	7600.2E6	195.614	212.415
11) T	2,4,5-TP ...	8.942	9.626	3037.8E6	2994.1E6	190.013	205.252
12) T	2,4,5-T	9.229	10.034	2459.5E6	2720.1E6	196.433	202.716
13) T	2,4-DB	9.792	10.594	331.8E6	292.1E6	189.139	199.965
14) T	DINOSEB	10.957	10.966	2238.2E6	2080.2E6	194.260	204.710
15) T	Picloram	10.780	12.017	2104.2E6	3281.5E6	164.298	160.179m
16) T	DCPA	11.257	11.991	3771.4E6	3977.5E6	192.980	205.680m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS033125\
Data File : PS029628.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2025 03:24
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/01/2025
Supervised By : mohammad ahmed 04/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 05:09:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS033125.M
Quant Title : 8080.M
QLast Update : Tue Apr 01 05:08:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

