

Data Path : P:\HPCHEM1\ECD_S\Data\PS041018\
 Data File : PS000109.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2018 16:56
 Operator : UA
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 07:27:38 2018
 Quant Method : P:\HPCHEM1\ECD_S\Method\PS041018.M
 Quant Title : 8080.M
 QLast Update : Wed Apr 11 07:22:06 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
10) T	Pentachlo...	14.109	13.631	6418235	5332425	1.296 1.282
11) T	2,4,5-TP ...	14.976	14.240	5410486	4272529	2.741 2.622
12) T	2,4,5-T	15.374	14.772	9196298	6541929	4.820 4.296
14) T	DINOSEB	17.643	15.894	7949879	6058765	5.790 5.395
15) T	Picloram	17.399	17.141	17191371	10875021	6.887 5.534
16) T	DCPA	17.983	17.083	17751079	13559021	7.598 7.016



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_S\Data\PS041018\
Data File : PS000109.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2018 16:56
Operator : UA
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 07:27:38 2018
Quant Method : P:\HPCHEM1\ECD_S\Method\PS041018.M
Quant Title : 8080.M
QLast Update : Wed Apr 11 07:22:06 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T.: 11.831 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T.: 11.200 min
Response: 0
Conc: N.D.




#10 Pentachlorophenol

R.T.: 14.109 min
Delta R.T.: 0.008 min
Response: 6418235
Conc: 1.30 ng/ml



#10 Pentachlorophenol


R.T.: 13.631 min
Delta R.T.: -0.019 min
Response: 5332425
Conc: 1.28 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.976 min
Delta R.T.: 0.009 min
Response: 5410486
Conc: 2.74 ng/ml

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 14.240 min
Delta R.T.: -0.017 min
Response: 4272529
Conc: 2.62 ng/ml




#12 2,4,5-T

R.T.: 15.374 min
Delta R.T.: 0.009 min
Response: 9196298
Conc: 4.82 ng/ml



#12 2,4,5-T


R.T.: 14.772 min
Delta R.T.: -0.016 min
Response: 6541929
Conc: 4.30 ng/ml



#14 DINOSEB


R.T.: 17.643 min
Delta R.T.: 0.009 min
Response: 7949879
Conc: 5.79 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 15.894 min
Delta R.T.: -0.015 min
Response: 6058765
Conc: 5.39 ng/ml




#15 Picloram

R.T.: 17.399 min
Delta R.T.: 0.010 min
Response: 17191371
Conc: 6.89 ng/ml



#15 Picloram


R.T.: 17.141 min
Delta R.T.: -0.014 min
Response: 10875021
Conc: 5.53 ng/ml



#16 DCPA

R.T.: 17.983 min
Delta R.T.: 0.009 min
Response: 17751079
Conc: 7.60 ng/ml

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 17.083 min
Delta R.T.: -0.015 min
Response: 13559021
Conc: 7.02 ng/ml