

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041119\
 Data File : PS004032.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2019 14:19
 Operator : EI\SJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS041119

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 14:47:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041119.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 11 14:18:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound		AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon		685.093	696.747 E3	-1.7	102	-0.04
2 T	3,5-DICHLORO BENZOIC ACID		715.626	717.661 E3	-0.3	96	-0.04
3 T	4-Nitrophenol		252.596	259.368 E3	-2.7	99	-0.03
4 S	2,4-DCAA		485.215	482.918 E3	0.5	101	-0.04
5 T	DICAMBA		1.962	1.995 E6	-1.7	100	-0.04
6 T	MCP P		1.305	1.350 E6	-3.4	100	-0.04
7 T	MCPA		2.063	2.110 E6	-2.3	101	-0.04
8 T	DICHLORPROP		484.099	496.270 E3	-2.5	101	-0.04
9 T	2,4-D		455.329	472.322 E3	-3.7	100	-0.04
10 T	Pentachlorophenol		7.253	7.490 E6	-3.3	100	-0.04
11 T	2,4,5-TP (SILVEX)		2.707	2.791 E6	-3.1	100	-0.04
12 T	2,4,5-T		2.944	2.997 E6	-1.8	99	-0.04
13 T	2,4-DB		311.419	328.740 E3	-5.6	97	-0.04
14 T	DINOSEB		2.505	2.491 E6	0.6	100	-0.04
15 T	Picloram		3.586	3.741 E6	-4.3	100	-0.04
16 T	DCPA		3.600	3.655 E6	-1.5	101	-0.04

Signal #2

1 T	Dalapon		628.454	647.447 E3	-3.0	102	-0.03
2 T	3,5-DICHLORO BENZOIC ACID		571.625	590.582 E3	-3.3	105	-0.03
3 T	4-Nitrophenol		239.807	253.036 E3	-5.5	101	-0.03
4 S	2,4-DCAA		459.242	475.845 E3	-3.6	102	-0.03
5 T	DICAMBA		1.646	1.687 E6	-2.5	100	-0.04
6 T	MCP P		1.043	1.074 E6	-3.0	99	-0.04
7 T	MCPA		1.820	1.832 E6	-0.7	103	-0.04
8 T	DICHLORPROP		424.517	429.099 E3	-1.1	100	-0.04
9 T	2,4-D		511.713	520.916 E3	-1.8	101	-0.04
10 T	Pentachlorophenol		6.210	6.410 E6	-3.2	100	-0.04
11 T	2,4,5-TP (SILVEX)		2.408	2.483 E6	-3.1	100	-0.04
12 T	2,4,5-T		2.307	2.423 E6	-5.0	101	-0.03
13 T	2,4-DB		308.086	329.167 E3	-6.8	101	-0.03
14 T	DINOSEB		1.869	1.925 E6	-3.0	100	-0.03
15 T	Picloram		3.423	3.587 E6	-4.8	101	-0.04
16 T	DCPA		3.178	3.295 E6	-3.7	101	-0.04

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041119\
Data File : PS004032.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2019 14:19
Operator : EI\SJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS041119

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 14:47:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041119.M
Quant Title : 8080.M
QLast Update : Thu Apr 11 14:18:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0