

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042219\
 Data File : PS004210.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2019 04:00
 Operator : EI\SJ
 Sample : K2475-08
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-TRACK-78-79

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 07:06:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041119.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 11 14:18:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.205	10.711	289.8E6	240.8E6	597.218	524.243
Target Compounds							
2) T	3,5-DICHL...	0.000	8.879	0	3133829	N.D.	5.482
3) T	4-Nitroph...	10.845	9.930	17569955	19177559	69.557	79.971
5) T	DICAMBA	11.562	0.000	15018530	0	7.656	N.D. #
6) T	MCP	11.882	0.000	18178443	0	13.925	N.D. #
7) T	MCPA	0.000	11.602	0	2219933	N.D.	1.220
8) T	DICHLORPROP	12.680	0.000	17186661	0	35.502	N.D. #
9) T	2,4-D	13.056	0.000	20224515	0	44.417	N.D. #
10) T	Pentachlo...	13.433	13.199	29130554	1980340	4.016	<MDL #
11) T	2,4,5-TP ...	14.315	13.715	19121215	604405	7.063	<MDL #
12) T	2,4,5-T	14.742	14.308	18060826	10172607	6.134	4.410 #
14) T	DINOSEB	16.924	15.450	11501698	5350297	4.591	2.863 #
15) T	Picloram	16.687	0.000	54169655	0	15.105	N.D. #
16) T	DCPA	17.339	0.000	31313898	0	8.699	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042219\
Data File : PS004210.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Apr 2019 04:00
Operator : EI\SJ
Sample : K2475-08
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TP-TRACK-78-79

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 07:06:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041119.M
Quant Title : 8080.M
QLast Update : Thu Apr 11 14:18:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 0.000 min
Exp R.T.: 9.673 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
TP-TRACK-78-79




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 8.879 min
Delta R.T.: -0.030 min
Response: 3133829
Conc: 5.48 ng/ml




#3 4-Nitrophenol

R.T.: 10.845 min
Delta R.T.: 0.002 min
Response: 17569955
Conc: 69.56 ng/ml



#3 4-Nitrophenol


R.T.: 9.930 min
Delta R.T.: 0.032 min
Response: 19177559
Conc: 79.97 ng/ml



#4 2,4-DCAA


R.T.: 11.205 min
Delta R.T.: -0.038 min
Response: 289778713
Conc: 597.22 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-TRACK-78-79




#4 2,4-DCAA

R.T.: 10.711 min
Delta R.T.: -0.035 min
Response: 240754559
Conc: 524.24 ng/ml




#5 DICAMBA

R.T.: 11.562 min
Delta R.T.: 0.020 min
Response: 15018530
Conc: 7.66 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 11.023 min
Response: 0
Conc: N.D.



#6 MCPP


R.T.: 11.882 min
Delta R.T.: 0.000 min
Response: 18178443
Conc: 13.92 ug/ml

Instrument :
ECD_S
ClientSampleId :
TP-TRACK-78-79




#6 MCPP

R.T.: 0.000 min
Exp R.T. : 11.255 min
Response: 0
Conc: N.D.




#7 MCPA

R.T.: 0.000 min
Exp R.T. : 12.132 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 11.602 min
Delta R.T.: -0.028 min
Response: 2219933
Conc: 1.22 ug/ml



#8 DICHLORPROP


R.T.: 12.680 min
Delta R.T.: -0.054 min
Response: 17186661
Conc: 35.50 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-TRACK-78-79




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 12.149 min
Response: 0
Conc: N.D.




#9 2,4-D

R.T.: 13.056 min
Delta R.T.: -0.046 min
Response: 20224515
Conc: 44.42 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T. : 12.623 min
Response: 0
Conc: N.D.



#10 Pentachlorophenol


R.T.: 13.433 min
Delta R.T.: -0.019 min
Response: 29130554
Conc: 4.02 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-TRACK-78-79




#10 Pentachlorophenol

R.T.: 13.199 min
Delta R.T.: 0.052 min
Response: 1980340
Conc: N.D.




#11 2,4,5-TP (SILVEX)

R.T.: 14.315 min
Delta R.T.: -0.018 min
Response: 19121215
Conc: 7.06 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 13.715 min
Delta R.T.: -0.063 min
Response: 604405
Conc: N.D.



#12 2,4,5-T


R.T.: 14.742 min
Delta R.T.: 0.006 min
Response: 18060826
Conc: 6.13 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-TRACK-78-79




#12 2,4,5-T

R.T.: 14.308 min
Delta R.T.: -0.030 min
Response: 10172607
Conc: 4.41 ng/ml




#14 DINOSEB

R.T.: 16.924 min
Delta R.T.: -0.012 min
Response: 11501698
Conc: 4.59 ng/ml



#14 DINOSEB


R.T.: 15.450 min
Delta R.T.: 0.039 min
Response: 5350297
Conc: 2.86 ng/ml



#15 Picloram


R.T.: 16.687 min
Delta R.T.: -0.034 min
Response: 54169655
Conc: 15.11 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-TRACK-78-79




#15 Picloram

R.T.: 0.000 min
Exp R.T. : 16.642 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 17.339 min
Delta R.T.: 0.056 min
Response: 31313898
Conc: 8.70 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 16.584 min
Response: 0
Conc: N.D.