

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042219\
 Data File : PS004206.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2019 01:51
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
ClientSampleId :
 HSTDCCC750

Manual Integrations
APPROVED
 mohammad
 4/25/2019 4:56:28 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 07:05:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041119.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 11 14:18:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.187	10.692	353.4E6	379.5E6	728.291m	826.439
Target Compounds							
1) T	Dalapon	2.810	2.295	506.8E6	454.9E6	739.826	723.795
2) T	3,5-DICHL...	9.616	8.855	550.7E6	429.5E6	769.590m	751.280m
3) T	4-Nitroph...	10.800	9.848	159.0E6	168.4E6	629.611m	702.164
5) T	DICAMBA	11.485	10.967	1603.1E6	1292.9E6	817.270	785.252m
6) T	MCPP	11.825	11.200	110.1E6	83913551	84.351	80.452
7) T	MCPA	12.075	11.575	168.7E6	131.4E6	81.762	72.239
8) T	DICHLORPROP	12.676	12.095	383.9E6	329.8E6	793.099m	776.785
9) T	2,4-D	13.046	12.570	345.6E6	361.5E6	758.950m	706.503m
10) T	Pentachlo...	13.394	13.093	5826.6E6	4867.8E6	803.336m	783.808m
11) T	2,4,5-TP ...	14.276	13.724	2215.1E6	1888.9E6	818.220	784.412
12) T	2,4,5-T	14.681	14.287	2447.0E6	1764.2E6	831.091	764.831
13) T	2,4-DB	15.426	14.941	233.1E6	220.1E6	748.491m	714.268
14) T	DINOSEB	16.878	15.359	1980.6E6	1368.1E6	790.530	732.012m
15) T	Picloram	16.667	16.592	2541.4E6	2239.6E6	708.659	654.200m
16) T	DCPA	17.227	16.531	2799.7E6	2503.9E6	777.753m	787.800m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042219\
 Data File : PS004206.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2019 01:51
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
Client Sampled :
 HSTDCCC750

Manual Integrations
APPROVED
 mohammad
 4/25/2019 4:56:28 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 07:05:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041119.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 11 14:18:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

