

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042220\
 Data File : PS009979.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2020 20:54
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 01:49:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.235	6.922	489.3E6	415.8E6	810.361	839.564
Target Compounds							
1) T	Dalapon	3.045	2.629	445.2E6	426.5E6	587.690	595.730
2) T	3,5-DICHL...	6.446	6.071	610.0E6	572.8E6	741.752	771.015
3) T	4-Nitroph...	7.064	6.525	210.1E6	227.7E6	755.315	749.900
5) T	DICAMBA	7.428	7.091	1750.5E6	1575.0E6	777.101	791.858
6) T	MCPD	7.598	7.190	139.2E6	130.3E6	77.982	79.169
7) T	MCPA	7.748	7.406	196.7E6	188.9E6	76.299	79.185
8) T	DICHLORPROP	8.121	7.717	420.7E6	406.4E6	758.024	807.479
9) T	2,4-D	8.350	8.011	444.7E6	475.2E6	774.446	793.009
10) T	Pentachlo...	8.663	8.443	5761.2E6	4902.7E6	775.306	785.685
11) T	2,4,5-TP ...	9.212	8.803	2347.4E6	2130.3E6	793.187	814.566
12) T	2,4,5-T	9.497	9.183	2220.4E6	2121.3E6	800.832	812.910
13) T	2,4-DB	10.050	9.689	303.0E6	287.4E6	799.772	810.452
14) T	DINOSEB	11.215	10.024	1669.8E6	1560.2E6	746.213	792.299
15) T	Picloram	11.029	11.003	2397.9E6	2409.2E6	773.995	708.350
16) T	DCPA	11.479	10.953	2898.5E6	2708.3E6	778.387	807.726

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042220\
Data File : PS009979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2020 20:54
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 01:49:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

