

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS042419\  
 Data File : PS004252.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 24 Apr 2019 12:31  
 Operator : EI\SJ  
 Sample : I.BLK  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 ECD\_S  
 ClientSampleId :  
 I.BLK

Manual Integrations  
 APPROVED

mohammad  
 4/26/2019 1:21:46 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Apr 25 00:51:42 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS041119.M  
 Quant Title : 8080.M  
 QLast Update : Thu Apr 11 14:18:42 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2   | Resp#1  | Resp#2  | ng/ml   | ng/ml    |
|-----------------------------|--------|--------|---------|---------|---------|----------|
| -----                       |        |        |         |         |         |          |
| System Monitoring Compounds |        |        |         |         |         |          |
| 4) S 2,4-DCAA               | 11.223 | 10.712 | 178.3E6 | 168.9E6 | 367.539 | 367.825m |

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS042419\  
Data File : PS004252.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 24 Apr 2019 12:31  
Operator : EI\SJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1



Instrument :  
ECD\_S  
ClientSampleId :  
I.BLK


Manual Integrations  
APPROVED

mohammad  
4/26/2019 1:21:46 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Apr 25 00:51:42 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS041119.M  
Quant Title : 8080.M  
QLast Update : Thu Apr 11 14:18:42 2019  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 11.223 min  
Delta R.T.: -0.020 min  
Response: 178335292  
Conc: 367.54 ng/ml

Instrument :  
ECD\_S  
ClientSampleId :  
I.BLK

Manual Integrations  
APPROVED

mohammad  
4/26/2019 1:21:46 PM



#4 2,4-DCAA

R.T.: 10.712 min  
Delta R.T.: -0.034 min  
Response: 168920579  
Conc: 367.82 ng/ml m