

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS043020\
 Data File : PS010079.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Apr 2020 12:39
 Operator : DD\AJ
 Sample : L2436-04
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 14:45:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.235	6.922	158.5E6	144.9E6	262.470	292.524
Target Compounds							
1) T	Dalapon	0.000	2.581	0	26015737	N.D.	36.340
2) T	3,5-DICHL...	0.000	6.052	0	3412795	N.D.	4.594
3) T	4-Nitroph...	7.009	6.512	9366661	1492996	33.669	4.916 #
5) T	DICAMBA	7.418	7.090	7178438	3818376	3.187	1.920 #
6) T	MCPD	0.000	7.209	0	1272908	N.D.	<MDL
7) T	MCPA	7.701	7.392	34377049	2694850	13.337	1.130 #
8) T	DICHLORPROP	0.000	7.740	0	9661186	N.D.	19.195
9) T	2,4-D	0.000	7.990	0	923919	N.D.	1.542
10) T	Pentachlo...	0.000	8.440	0	7553434	N.D.	1.210
11) T	2,4,5-TP ...	9.249	8.845	15027332	39501791	5.078	15.105 #
12) T	2,4,5-T	9.441	9.184	73062457	6359662	26.351	2.437 #
13) T	2,4-DB	10.089	9.705	15415225	2875575	40.694	8.110 #
14) T	DINOSEB	0.000	10.029	0	10237149	N.D.	5.198
15) T	Picloram	11.002	11.005	68638942	19935963	22.155	5.861 #
16) T	DCPA	0.000	10.952	0	8640702	N.D.	2.577



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS043020\
Data File : PS010079.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Apr 2020 12:39
Operator : DD\AJ
Sample : L2436-04
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 14:45:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 3.040 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.581 min
Delta R.T.: -0.040 min
Response: 26015737
Conc: 36.34 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T.: 6.448 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.052 min
Delta R.T.: -0.022 min
Response: 3412795
Conc: 4.59 ng/ml



#3 4-Nitrophenol


R.T.: 7.009 min
Delta R.T.: -0.059 min
Response: 9366661
Conc: 33.67 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.512 min
Delta R.T.: -0.019 min
Response: 1492996
Conc: 4.92 ng/ml




#4 2,4-DCAA

R.T.: 7.235 min
Delta R.T.: -0.004 min
Response: 158483359
Conc: 262.47 ng/ml



#4 2,4-DCAA


R.T.: 6.922 min
Delta R.T.: -0.007 min
Response: 144872639
Conc: 292.52 ng/ml



#5 DICAMBA


R.T.: 7.418 min
Delta R.T.: -0.014 min
Response: 7178438
Conc: 3.19 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 7.090 min
Delta R.T.: -0.008 min
Response: 3818376
Conc: 1.92 ng/ml




#7 MCPA

R.T.: 7.701 min
Delta R.T.: -0.050 min
Response: 34377049
Conc: 13.34 ug/ml



#7 MCPA


R.T.: 7.392 min
Delta R.T.: -0.019 min
Response: 2694850
Conc: 1.13 ug/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 8.125 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 7.740 min
Delta R.T.: 0.015 min
Response: 9661186
Conc: 19.20 ng/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T.: 8.354 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 7.990 min
Delta R.T.: -0.029 min
Response: 923919
Conc: 1.54 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T.: 8.667 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 8.440 min
Delta R.T.: -0.012 min
Response: 7553434
Conc: 1.21 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.249 min
Delta R.T.: 0.033 min
Response: 15027332
Conc: 5.08 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 8.845 min
Delta R.T.: 0.033 min
Response: 39501791
Conc: 15.10 ng/ml



#12 2,4,5-T


R.T.: 9.441 min
Delta R.T.: -0.059 min
Response: 73062457
Conc: 26.35 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 9.184 min
Delta R.T.: -0.008 min
Response: 6359662
Conc: 2.44 ng/ml




#13 2,4-DB

R.T.: 10.089 min
Delta R.T.: 0.036 min
Response: 15415225
Conc: 40.69 ng/ml



#13 2,4-DB


R.T.: 9.705 min
Delta R.T.: 0.006 min
Response: 2875575
Conc: 8.11 ng/ml



#14 DINOSEB


R.T.: 0.000 min
Exp R.T.: 11.219 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 10.029 min
Delta R.T.: -0.006 min
Response: 10237149
Conc: 5.20 ng/ml




#15 Picloram

R.T.: 11.002 min
Delta R.T.: -0.030 min
Response: 68638942
Conc: 22.16 ng/ml



#15 Picloram


R.T.: 11.005 min
Delta R.T.: -0.008 min
Response: 19935963
Conc: 5.86 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 11.484 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 10.952 min
Delta R.T.: -0.012 min
Response: 8640702
Conc: 2.58 ng/ml