

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052019\
 Data File : PS004660.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 May 2019 17:40
 Operator : EI\SJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Manual Integrations
 APPROVED

Ankita
 5/21/2019 4:50:50 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 21 04:41:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:36:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.594	5.303	1311.4E6	955.0E6	945.657	987.785
Target Compounds							
1) T	Dalapon	2.117	1.813	1526.7E6	1218.6E6	907.733m	968.524
2) T	3,5-DICHL...	4.941	4.595	1561.9E6	1123.0E6	885.078	912.347
3) T	4-Nitroph...	5.439	4.977	566.0E6	430.4E6	899.596	903.299
5) T	DICAMBA	5.754	5.438	4718.4E6	3372.9E6	883.280	911.164
6) T	MCPD	5.916	5.529	387.8E6	262.5E6	91.101	94.624
7) T	MCPA	6.055	5.715	528.9E6	377.9E6	89.008	92.599
8) T	DICHLORPROP	6.394	5.994	1161.6E6	873.0E6	879.926	930.576
9) T	2,4-D	6.609	6.266	1231.1E6	1000.8E6	887.032	928.876
10) T	Pentachlo...	6.881	6.644	15162.8E6	10830.9E6	864.400	890.005
11) T	2,4,5-TP ...	7.419	7.006	6413.1E6	4728.0E6	865.258	924.897
12) T	2,4,5-T	7.692	7.404	6552.0E6	4056.1E6	867.238	895.865m
13) T	2,4-DB	8.225	7.856	888.1E6	623.2E6	866.037	946.204
14) T	DINOSEB	9.328	8.173	5161.9E6	3538.7E6	861.145	919.581
15) T	Picloram	9.152	9.112	9548.3E6	6959.1E6	888.712	918.702
16) T	DCPA	9.599	9.073	7739.5E6	5921.7E6	853.143	919.031

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052019\
Data File : PS004660.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 May 2019 17:40
Operator : EI\SJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Manual Integrations
APPROVED

Ankita
5/21/2019 4:50:50 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 04:41:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:36:59 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

