

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS053019\
 Data File : PS004831.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2019 11:50
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 5/31/2019 8:28:23 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 02:16:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:53:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.576	5.284	1086.2E6	746.1E6	774.891	773.253
Target Compounds							
1) T	Dalapon	2.102	1.795	1224.0E6	991.3E6	738.379	768.169m
2) T	3,5-DICHL...	4.923	4.575	1296.6E6	890.6E6	751.858	722.793
3) T	4-Nitroph...	5.420	4.957	484.9E6	343.7E6	777.751	725.320
5) T	DICAMBA	5.738	5.418	3936.1E6	2682.3E6	741.314	726.089
6) T	MCPD	5.899	5.509	310.8E6	198.6E6	74.647	72.485
7) T	MCPA	6.038	5.694	438.1E6	284.2E6	75.628	69.464
8) T	DICHLORPROP	6.377	5.974	955.7E6	685.5E6	730.236	727.030
9) T	2,4-D	6.592	6.246	1081.3E6	794.4E6	790.461	734.993
10) T	Pentachlo...	6.866	6.625	13001.4E6	8866.4E6	740.364	721.736
11) T	2,4,5-TP ...	7.403	6.987	5444.0E6	3760.7E6	743.156	730.892
12) T	2,4,5-T	7.675	7.384	5428.8E6	3256.5E6	728.945	730.679
13) T	2,4-DB	8.207	7.836	779.7E6	493.0E6	807.217	754.231
14) T	DINOSEB	9.312	8.153	4046.9E6	2707.7E6	684.185	704.361
15) T	Picloram	9.134	9.093	7897.9E6	5460.8E6	757.954	731.957
16) T	DCPA	9.582	9.054	6418.2E6	4754.3E6	720.656	733.905

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS053019\
Data File : PS004831.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2019 11:50
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
5/31/2019 8:28:23 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 02:16:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:53:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

