

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS053019\
 Data File : PS004843.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2019 20:23
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 5/31/2019 8:28:39 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 02:21:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:53:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.572	5.284	1142.3E6	760.2E6	814.888	787.941
Target Compounds							
1) T	Dalapon	2.100	1.796	1187.7E6	935.1E6	716.484	724.642m
2) T	3,5-DICHL...	4.921	4.576	1362.0E6	907.9E6	789.769	736.878
3) T	4-Nitroph...	5.416	4.958	431.2E6	348.5E6	691.635m	735.304
5) T	DICAMBA	5.734	5.418	4141.6E6	2720.1E6	780.010	736.322
6) T	MCPD	5.895	5.509	327.1E6	209.6E6	78.563	76.500
7) T	MCPA	6.033	5.695	455.6E6	299.0E6	78.654	73.078
8) T	DICHLORPROP	6.373	5.974	991.0E6	698.9E6	757.221	741.233
9) T	2,4-D	6.587	6.245	1101.6E6	807.2E6	805.339	746.772
10) T	Pentachlo...	6.861	6.624	13701.3E6	9025.0E6	780.223	734.646
11) T	2,4,5-TP ...	7.398	6.986	5765.0E6	3827.0E6	786.969	743.792
12) T	2,4,5-T	7.670	7.384	5743.5E6	3210.4E6	771.200	720.348
13) T	2,4-DB	8.202	7.836	779.2E6	503.4E6	806.630m	770.174
14) T	DINOSEB	9.306	8.152	4427.3E6	2802.0E6	748.489	728.893
15) T	Picloram	9.130	9.092	7237.2E6	4834.7E6	694.550	648.039
16) T	DCPA	9.577	9.054	7010.3E6	4884.1E6	787.129	753.936

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS053019\
Data File : PS004843.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2019 20:23
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
5/31/2019 8:28:39 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 02:21:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:53:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

