

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060422\
 Data File : PS019616.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2022 20:25
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 22:52:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.197	6.782	2024.3E6	671.6E6	843.747	866.703
Target Compounds							
1) T	Dalapon	3.014	2.524	3409.8E6	748.0E6	672.700	745.405
2) T	3,5-DICHL...	6.411	5.962	2470.6E6	978.4E6	781.670	816.257
3) T	4-Nitroph...	7.029	6.400	849.2E6	341.7E6	865.381	798.267
5) T	DICAMBA	7.387	6.947	7507.8E6	2831.7E6	790.483	828.352
6) T	MCPD	7.557	7.045	612.0E6	226.5E6	81.535	83.660
7) T	MCPA	7.706	7.258	858.2E6	313.9E6	79.434	81.680
8) T	DICHLORPROP	8.078	7.566	2001.0E6	727.9E6	781.468	823.976
9) T	2,4-D	8.307	7.858	2335.4E6	876.5E6	816.710	821.745
10) T	Pentachlo...	8.613	8.283	27036.1E6	9470.1E6	781.640	812.449
11) T	2,4,5-TP ...	9.165	8.644	12361.4E6	4168.9E6	787.302	829.281
12) T	2,4,5-T	9.449	9.021	14706.3E6	3925.9E6	781.434	834.718
13) T	2,4-DB	10.002	9.526	3369.0E6	521.4E6	832.675	843.742
14) T	DINOSEB	11.155	9.851	12804.1E6	2980.7E6	827.823	839.012
15) T	Picloram	10.979	10.833	23942.5E6	5066.8E6	806.763	787.721
16) T	DCPA	11.423	10.781	19905.8E6	5117.2E6	825.828	829.035

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060422\
Data File : PS019616.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2022 20:25
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 22:52:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

