

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060519\
 Data File : PS004935.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2019 15:41
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 6/6/2019 11:03:25 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 01:02:21 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:53:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.574	5.286	1047.8E6	756.8E6	747.479	784.357
Target Compounds							
1) T	Dalapon	2.103	1.799	1144.6E6	822.7E6	690.452	637.493
2) T	3,5-DICHL...	4.922	4.578	1256.9E6	900.1E6	728.826	730.547
3) T	4-Nitroph...	5.418	4.960	454.7E6	340.3E6	729.192	718.066
5) T	DICAMBA	5.735	5.420	3809.2E6	2695.4E6	717.402	729.611
6) T	MCPD	5.897	5.511	296.3E6	207.9E6	71.185	75.869
7) T	MCPA	6.034	5.696	418.4E6	297.4E6	72.227	72.688
8) T	DICHLORPROP	6.374	5.976	915.9E6	692.0E6	699.864	734.013
9) T	2,4-D	6.588	6.248	1001.4E6	793.9E6	732.087	734.468
10) T	Pentachlo...	6.862	6.627	12814.6E6	9043.2E6	729.729	736.123
11) T	2,4,5-TP ...	7.398	6.988	5261.4E6	3813.5E6	718.229	741.151m
12) T	2,4,5-T	7.670	7.386	5237.4E6	3262.6E6	703.241	732.050
13) T	2,4-DB	8.203	7.838	735.5E6	496.1E6	761.391	758.986
14) T	DINOSEB	9.307	8.154	3961.6E6	2740.9E6	669.767	713.007
15) T	Picloram	9.130	9.095	7420.5E6	5401.1E6	712.140	723.960
16) T	DCPA	9.578	9.055	6328.7E6	4846.7E6	710.601	748.164

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060519\
Data File : PS004935.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2019 15:41
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
6/6/2019 11:03:25 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 01:02:21 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:53:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

