

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060619\
 Data File : PS004966.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2019 20:28
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 16:51:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:53:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.574	5.287	1154.2E6	783.9E6	823.416	812.507
Target Compounds							
1) T	Dalapon	2.104	1.802	1194.2E6	891.2E6	720.390	690.570
2) T	3,5-DICHL...	4.923	4.580	1621.1E6	942.7E6	940.027	765.103
3) T	4-Nitroph...	5.419	4.961	500.4E6	353.9E6	802.612	746.784
5) T	DICAMBA	5.736	5.422	4156.5E6	3000.1E6	782.817	812.099
6) T	MCPD	5.898	5.512	330.3E6	212.9E6	79.348	77.694
7) T	MCPA	6.035	5.698	461.7E6	310.0E6	79.710	75.766
8) T	DICHLORPROP	6.375	5.978	1014.6E6	731.6E6	775.276	775.971
9) T	2,4-D	6.589	6.249	1436.7E6	826.4E6	1050.305	764.537 #
10) T	Pentachlo...	6.863	6.628	13784.5E6	9312.2E6	784.958	758.022
11) T	2,4,5-TP ...	7.399	6.990	5825.6E6	3964.8E6	795.245	770.569
12) T	2,4,5-T	7.671	7.387	5764.1E6	3310.2E6	773.957	742.736
13) T	2,4-DB	8.203	7.840	816.0E6	517.3E6	844.785	791.431
14) T	DINOSEB	9.309	8.156	4418.9E6	2866.4E6	747.069	745.647
15) T	Picloram	9.131	9.096	7616.6E6	5071.0E6	730.962	679.716
16) T	DCPA	9.579	9.057	7110.9E6	5057.0E6	798.434	780.636

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060619\
Data File : PS004966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2019 20:28
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 16:51:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:53:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

