

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060622\
 Data File : PS019623.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2022 16:35
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 01:15:13 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.198	6.783	1785.0E6	587.8E6	744.007	758.535
Target Compounds							
1) T	Dalapon	3.013	2.525	3466.5E6	725.2E6	683.902	722.724
2) T	3,5-DICHL...	6.412	5.963	2196.8E6	866.0E6	695.062	722.446
3) T	4-Nitroph...	7.030	6.401	726.2E6	295.8E6	740.057	691.165
5) T	DICAMBA	7.387	6.947	6652.4E6	2503.3E6	700.421	732.282
6) T	MCPD	7.557	7.044	534.0E6	197.2E6	71.147	72.842
7) T	MCPA	7.705	7.257	748.2E6	274.0E6	69.252	71.300
8) T	DICHLORPROP	8.078	7.566	1764.2E6	637.2E6	688.980	721.307
9) T	2,4-D	8.308	7.859	2044.9E6	763.8E6	715.128	716.114
10) T	Pentachlo...	8.614	8.283	24466.4E6	8569.7E6	707.348	735.204
11) T	2,4,5-TP ...	9.165	8.644	11235.0E6	3699.2E6	715.564	735.838
12) T	2,4,5-T	9.450	9.022	13308.8E6	3459.1E6	707.176	735.482
13) T	2,4-DB	10.002	9.527	2968.5E6	450.4E6	733.689	728.819
14) T	DINOSEB	11.155	9.851	11518.9E6	2634.6E6	744.730	741.595
15) T	Picloram	10.980	10.833	22026.8E6	4596.6E6	742.210	714.624
16) T	DCPA	11.423	10.781	17880.8E6	4548.9E6	741.819	736.963

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060622\
Data File : PS019623.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2022 16:35
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 01:15:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

