

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS061219.M
 Title : 8080.M
 Last Update : Fri Jun 14 09:57:49 2019
 Response Via : Initial Calibration

Calibration Files

750 =PS005012.D 1000 =PS005013.D 1500 =PS005014.D
 500 =PS005011.D 200 =PS005010.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	1.734	1.757	1.770	1.708	1.739	1.742 E6	1.36
2) T	3,5-DICHLOROBENZOI...	1.859	1.851	1.804	1.932	1.978	1.885 E6	3.67
3) T	4-Nitrophenol	4.581	5.051	4.957	4.127	4.058	4.555 E5	10.05
4) S	2,4-DCAA	1.305	1.325	1.312	1.324	1.322	1.318 E6	0.67
5) T	DICAMBA	5.571	5.496	5.255	5.836	6.140	5.660 E6	5.99
6) T	MCPP	4.422	4.545	4.643	4.417	4.275	4.460 E6	3.14
7) T	MCPA	6.231	6.342	6.210	6.311	6.241	6.267 E6	0.91
8) T	DICHLORPROP	1.372	1.377	1.350	1.411	1.486	1.399 E6	3.80
9) T	2,4-D	1.348	1.377	1.387	1.357	1.278	1.349 E6	3.17
10) T	Pentachlorophenol	1.878	1.786	1.615	2.040	2.300	1.924 E7	13.55
11) T	2,4,5-TP (SILVEX)	7.716	7.543	7.186	8.137	8.705	7.857 E6	7.44
12) T	2,4,5-T	7.803	7.649	7.276	8.223	9.001	7.990 E6	8.25
13) T	2,4-DB	9.340	9.714	9.972	8.798	8.075	9.180 E5	8.27
14) T	DINOSEB	6.365	6.205	5.942	6.746	7.528	6.557 E6	9.39
15) T	Picloram	1.006	1.023	1.011	1.005	0.919	0.993 E7	4.22
16) T	DCPA	0.947	0.918	0.865	1.011	1.114	0.971 E7	9.84

Signal #2 Calibration Files

750 =PS005012.D 1000 =PS005013.D 1500 =PS005014.D
 500 =PS005011.D 200 =PS005010.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	1.333	1.352	1.302	1.263	1.277	1.306 E6	2.84
2) T	3,5-DICHLOROBENZOI...	1.253	1.245	1.214	1.304	1.384	1.280 E6	5.18
3) T	4-Nitrophenol	4.479	4.557	4.674	4.514	4.449	4.535 E5	1.93
4) S	2,4-DCAA	0.971	0.974	0.964	1.006	1.072	0.998 E6	4.50
5) T	DICAMBA	3.778	3.704	3.511	3.973	4.237	3.840 E6	7.20
6) T	MCPP	2.848	2.903	2.927	2.903	2.866	2.889 E6	1.10
7) T	MCPA	4.095	4.131	4.133	4.239	4.480	4.215 E6	3.73
8) T	DICHLORPROP	0.950	0.951	0.939	0.978	1.038	0.971 E6	4.13
9) T	2,4-D	1.049	1.066	1.052	1.076	1.108	1.070 E6	2.20
10) T	Pentachlorophenol	1.254	1.182	1.060	1.379	1.567	1.288 E7	15.05
11) T	2,4,5-TP (SILVEX)	5.241	5.124	4.843	5.570	5.986	5.353 E6	8.21
12) T	2,4,5-T	4.703	4.682	4.529	4.846	4.823	4.717 E6	2.70
13) T	2,4-DB	6.251	6.508	6.578	6.252	6.143	6.346 E5	2.93
14) T	DINOSEB	3.829	3.789	3.695	3.965	4.145	3.885 E6	4.51
15) T	Picloram	6.831	6.857	6.641	6.973	6.805	6.821 E6	1.75
16) T	DCPA	6.455	6.255	5.867	6.871	7.526	6.594 E6	9.61

(#) = Out of Range