

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061219\
 Data File : PS005013.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2019 02:51
 Operator : SM\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 07:56:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061219.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 13 07:54:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.586	5.299	1325.5E6	974.1E6	1015.667	1003.057
Target Compounds							
1) T	Dalapon	2.112	1.810	1599.3E6	1230.2E6	922.096	922.683
2) T	3,5-DICHL...	4.935	4.592	1721.8E6	1157.7E6	926.409	924.258
3) T	4-Nitroph...	5.434	4.975	459.7E6	414.7E6	1003.427	925.703
5) T	DICAMBA	5.746	5.433	5166.6E6	3481.9E6	927.464	921.650
6) T	MCPD	5.909	5.525	427.3E6	272.9E6	96.628	95.796
7) T	MCPA	6.046	5.711	589.8E6	384.2E6	94.666	93.834
8) T	DICHLORPROP	6.386	5.989	1294.0E6	893.9E6	943.510	940.960
9) T	2,4-D	6.601	6.261	1294.4E6	1002.2E6	960.240	955.378
10) T	Pentachlo...	6.872	6.638	16966.1E6	11232.6E6	903.420	895.640
11) T	2,4,5-TP ...	7.409	7.000	7165.4E6	4867.4E6	928.700	928.658
12) T	2,4,5-T	7.683	7.398	7266.8E6	4237.7E6	931.330	929.347
13) T	2,4-DB	8.216	7.851	922.8E6	618.2E6	987.994	988.992
14) T	DINOSEB	9.316	8.166	5833.1E6	3561.4E6	916.438	930.207
15) T	Picloram	9.142	9.107	9717.9E6	6513.7E6	965.795	953.597
16) T	DCPA	9.587	9.066	8815.5E6	6004.7E6	931.129	930.280

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061219\
Data File : PS005013.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2019 02:51
Operator : SM\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 07:56:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061219.M
Quant Title : 8080.M
QLast Update : Thu Jun 13 07:54:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

