

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061418\
 Data File : PS000345.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jun 2018 13:20
 Operator : UA
 Sample : J3458-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 00:55:09 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052118.M
 Quant Title : 8080.M
 QLast Update : Thu May 31 04:46:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.814	11.186	93064630	94178999	263.582	299.836
Target Compounds							
1) T	Dalapon	3.108	0.000	9435714	0	19.739	N.D. #
3) T	4-Nitroph...	11.381	10.343	7565972	6472879	32.482	34.757
5) T	DICAMBA	0.000	11.429	0	3081681	N.D.	2.785 #
6) T	MCP	12.490	0.000	331590	0	0.297	N.D. #
7) T	MCPA	12.788	12.048	12538623	15481691	6.691	10.427 #
8) T	DICHLORPROP	13.311	0.000	2310544	0	6.696	N.D. #
10) T	Pentachlo...	14.063	0.000	10849023	0	2.171	N.D. #
14) T	DINOSEB	17.588	15.891	7799834	14564272	5.330	11.775 #
15) T	Picloram	17.417	17.142	5532776	10255415	2.115	4.813 #
16) T	DCPA	17.972	17.142	-5760763	10255415	N.D.	4.962



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061418\
Data File : PS000345.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jun 2018 13:20
Operator : UA
Sample : J3458-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 00:55:09 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052118.M
Quant Title : 8080.M
QLast Update : Thu May 31 04:46:50 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.108 min
Delta R.T.: -0.062 min
Response: 9435714
Conc: 19.74 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 0.000 min
Exp R.T. : 2.546 min
Response: 0
Conc: N.D.




#3 4-Nitrophenol

R.T.: 11.381 min
Delta R.T.: -0.044 min
Response: 7565972
Conc: 32.48 ng/ml



#3 4-Nitrophenol


R.T.: 10.343 min
Delta R.T.: 0.004 min
Response: 6472879
Conc: 34.76 ng/ml



#4 2,4-DCAA


R.T.: 11.814 min
Delta R.T.: -0.013 min
Response: 93064630
Conc: 263.58 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 11.186 min
Delta R.T.: -0.005 min
Response: 94178999
Conc: 299.84 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 12.147 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 11.429 min
Delta R.T.: -0.052 min
Response: 3081681
Conc: 2.79 ng/ml



#6 MCPP


R.T.: 12.490 min
Delta R.T.: 0.009 min
Response: 331590
Conc: 0.30 ug/ml

Instrument :
ECD_S
ClientSampleId :




#6 MCPP

R.T.: 0.000 min
Exp R.T. : 11.700 min
Response: 0
Conc: N.D.




#7 MCPA

R.T.: 12.788 min
Delta R.T.: 0.055 min
Response: 12538623
Conc: 6.69 ug/ml



#7 MCPA


R.T.: 12.048 min
Delta R.T.: -0.028 min
Response: 15481691
Conc: 10.43 ug/ml



#8 DICHLORPROP


R.T.: 13.311 min
Delta R.T.: -0.030 min
Response: 2310544
Conc: 6.70 ng/ml

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 12.605 min
Response: 0
Conc: N.D.




#10 Pentachlorophenol

R.T.: 14.063 min
Delta R.T.: -0.034 min
Response: 10849023
Conc: 2.17 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T. : 13.641 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 17.588 min
Delta R.T.: -0.043 min
Response: 7799834
Conc: 5.33 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 15.891 min
Delta R.T.: -0.011 min
Response: 14564272
Conc: 11.77 ng/ml




#15 Picloram

R.T.: 17.417 min
Delta R.T.: 0.030 min
Response: 5532776
Conc: 2.12 ng/ml



#15 Picloram


R.T.: 17.142 min
Delta R.T.: -0.007 min
Response: 10255415
Conc: 4.81 ng/ml



#16 DCPA

R.T.: 17.972 min
Delta R.T.: 0.000 min
Response: -5760763
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 17.142 min
Delta R.T.: 0.053 min
Response: 10255415
Conc: 4.96 ng/ml