

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061418\
 Data File : PS000361.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jun 2018 22:35
 Operator : UA
 Sample : J3459-02MSD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-RO-341MSD

Manual Integrations
 APPROVED

UGO
 6/15/2018 12:09:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 01:01:26 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052118.M
 Quant Title : 8080.M
 QLast Update : Thu May 31 04:46:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.824	11.198	122.2E6	113.8E6	346.019	362.161
Target Compounds							
1) T	Dalapon	3.176	2.556	153.4E6	122.6E6	320.960m	310.361
2) T	3,5-DICHL...	10.238	9.342	220.2E6	186.7E6	411.586	422.846
3) T	4-Nitroph...	11.483	10.354	4133392	5116489	17.745	27.474 #
5) T	DICAMBA	12.143	11.487	582.4E6	522.0E6	445.297	471.811
6) T	MCPD	12.479	11.709	47211066	41697912	42.290m	45.053m
7) T	MCPA	12.730	12.086	66075822	50636096	35.261	34.104
8) T	DICHLORPROP	13.335	12.611	165.8E6	140.0E6	480.555	484.471
9) T	2,4-D	13.693	13.082	211.4E6	198.6E6	468.781	516.516
10) T	Pentachlo...	14.092	13.646	1968.2E6	1758.3E6	393.790	399.131
11) T	2,4,5-TP ...	14.957	14.253	909.4E6	832.1E6	458.819	483.781
12) T	2,4,5-T	15.355	14.786	884.5E6	791.2E6	451.114	480.041
13) T	2,4-DB	16.109	15.472	126.2E6	108.7E6	464.046	512.242
14) T	DINOSEB	17.620	15.906	354.1E6	308.5E6	241.958	249.435
15) T	Picloram	17.376	17.151	1093.1E6	950.8E6	417.887	446.250
16) T	DCPA	17.960	17.093	796.1E6	725.4E6	328.817	350.990

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061418\
Data File : PS000361.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jun 2018 22:35
Operator : UA
Sample : J3459-02MSD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MH-RO-341MSD

Manual Integrations
APPROVED

UGO
6/15/2018 12:09:31 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 01:01:26 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052118.M
Quant Title : 8080.M
QLast Update : Thu May 31 04:46:50 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

