

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061418\
 Data File : PS000366.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2018 01:10
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

UGO
 6/15/2018 12:09:46 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 01:50:41 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052118.M
 Quant Title : 8080.M
 QLast Update : Thu May 31 04:46:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.846	11.220	274.0E6	251.4E6	776.085m	800.354m
Target Compounds							
1) T	Dalapon	3.199	2.573	326.7E6	274.4E6	683.337m	694.615m
2) T	3,5-DICHL...	10.261	9.365	371.4E6	321.6E6	694.339	728.426
3) T	4-Nitroph...	11.444	10.367	162.0E6	137.9E6	695.393	740.409
5) T	DICAMBA	12.166	11.509	983.0E6	873.9E6	751.581	789.895m
6) T	MCPD	12.504	11.735	81303625	68609032	72.829m	74.129m
7) T	MCPA	12.757	12.113	121.6E6	104.9E6	64.905	70.667m
8) T	DICHLORPROP	13.358	12.633	251.6E6	223.2E6	729.239	772.491
9) T	2,4-D	13.716	13.104	338.1E6	296.6E6	749.575	771.577
10) T	Pentachlo...	14.115	13.668	3684.3E6	3375.0E6	737.134	766.104
11) T	2,4,5-TP ...	14.980	14.276	1508.5E6	1385.1E6	761.061	805.235
12) T	2,4,5-T	15.378	14.808	1493.4E6	1349.5E6	761.615	818.794
13) T	2,4-DB	16.130	15.494	205.8E6	165.1E6	756.822	778.464
14) T	DINOSEB	17.642	15.927	1015.0E6	912.2E6	693.554	737.460m
15) T	Picloram	17.398	17.173	1624.9E6	1397.4E6	621.225	655.874
16) T	DCPA	17.981	17.114	1820.1E6	1675.4E6	751.735	810.710

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061418\
Data File : PS000366.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2018 01:10
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleID :
HSTDCCC750

Manual Integrations
APPROVED

UGO
6/15/2018 12:09:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 01:50:41 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052118.M
Quant Title : 8080.M
QLast Update : Thu May 31 04:46:50 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

