

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061518\
 Data File : PS000378.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2018 15:20
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

ugo
 6/18/2018 1:21:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 22:51:15 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052118.M
 Quant Title : 8080.M
 QLast Update : Thu May 31 04:46:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.808	11.181	273.6E6	253.3E6	775.040	806.292
Target Compounds							
1) T	Dalapon	3.160	2.538	329.2E6	286.5E6	688.719m	725.373
2) T	3,5-DICHL...	10.222	9.326	372.3E6	320.6E6	696.110	726.094
3) T	4-Nitroph...	11.406	10.329	162.4E6	136.1E6	696.999	730.879
5) T	DICAMBA	12.128	11.471	982.1E6	868.8E6	750.899	785.342
6) T	MCPD	12.467	11.697	81055942	67975002	72.607	73.444
7) T	MCPA	12.718	12.075	123.1E6	102.6E6	65.706	69.069
8) T	DICHLORPROP	13.320	12.595	255.4E6	221.6E6	740.281	767.152
9) T	2,4-D	13.677	13.065	336.0E6	300.6E6	745.001	781.985
10) T	Pentachlo...	14.077	13.630	3679.9E6	3376.1E6	736.248	766.353
11) T	2,4,5-TP ...	14.942	14.238	1510.9E6	1377.6E6	762.275	800.906
12) T	2,4,5-T	15.340	14.770	1484.0E6	1333.9E6	756.824	809.304
13) T	2,4-DB	16.093	15.456	206.3E6	172.0E6	758.660	810.670
14) T	DINOSEB	17.605	15.890	1008.6E6	900.7E6	689.243	728.206
15) T	Picloram	17.361	17.135	1535.5E6	1303.4E6	587.047	611.767
16) T	DCPA	17.943	17.077	1817.3E6	1666.5E6	750.604	806.371

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061518\
Data File : PS000378.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2018 15:20
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

ugo
6/18/2018 1:21:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 22:51:15 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052118.M
Quant Title : 8080.M
QLast Update : Thu May 31 04:46:50 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

