

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062119\
 Data File : PS005162.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2019 13:53
 Operator : SM\AJ
 Sample : PB120788BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB120788BS

Manual Integrations
 APPROVED

Ankita
 6/24/2019 8:59:15 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 22:56:28 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.584	5.298	916.4E6	625.9E6	624.300	624.452
Target Compounds							
1) T	Dalapon	2.113	1.813	1159.3E6	671.1E6	628.901	508.738
2) T	3,5-DICHL...	4.934	4.592	1132.3E6	744.0E6	563.413	564.215
3) T	4-Nitroph...	5.430	4.973	313.6E6	245.9E6	432.970	490.363
5) T	DICAMBA	5.744	5.432	2944.9E6	1991.0E6	516.227	531.884
6) T	MCPD	5.903	5.520	260.5E6	157.3E6	59.503	56.482
7) T	MCPA	6.040	5.704	332.7E6	244.9E6	52.977	60.447
8) T	DICHLORPROP	6.383	5.987	780.6E6	575.2E6	545.865	595.748
9) T	2,4-D	6.597	6.258	936.4E6	719.2E6	602.358	644.848
10) T	Pentachlo...	6.870	6.636	11543.3E6	7639.0E6	586.666	594.186
11) T	2,4,5-TP ...	7.407	6.998	4679.8E6	3133.9E6	573.630	588.495
12) T	2,4,5-T	7.679	7.360	4507.7E6	2878.5E6	560.111	564.315
13) T	2,4-DB	8.213	7.849	703.5E6	482.0E6	605.618	738.281m
14) T	DINOSEB	9.314	8.164	3356.2E6	2116.8E6	525.093	523.550
15) T	Picloram	9.140	9.104	5661.8E6	3648.3E6	492.709	483.441m
16) T	DCPA	9.585	9.064	3746.3E6	2151.9E6	375.560	316.117m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062119\
Data File : PS005162.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2019 13:53
Operator : SM\AJ
Sample : PB120788BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB120788BS

Manual Integrations
APPROVED

Ankita
6/24/2019 8:59:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 22:56:28 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

