

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062519\
 Data File : PS005207.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2019 20:30
 Operator : SM\AJ
 Sample : K3482-02MS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleID :
 TP-1MS

Manual Integrations
 APPROVED

Ankita
 6/26/2019 1:10:49 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 01:09:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.568	5.284	388.1E6	256.9E6	264.365m	256.258
Target Compounds							
1) T	Dalapon	2.096	1.799	611.0E6	395.3E6	331.448m	299.696m
2) T	3,5-DICHL...	4.918	4.576	715.8E6	483.0E6	356.185	366.271m
5) T	DICAMBA	5.728	5.417	2083.2E6	1423.3E6	365.165m	380.232
6) T	MCPD	5.887	5.505	314.3E6	140.4E6	71.798	50.395 #
7) T	MCPA	6.023	5.688	278.9E6	188.6E6	44.405m	46.562m
8) T	DICHLORPROP	6.367	5.971	650.8E6	465.1E6	455.101m	481.722m
9) T	2,4-D	6.581	6.243	864.6E6	744.5E6	556.176	667.561
10) T	Pentachlo...	6.854	6.622	6439.2E6	4221.4E6	327.258	328.353
11) T	2,4,5-TP ...	7.390	6.983	4031.8E6	2935.1E6	494.207	551.163
12) T	2,4,5-T	7.662	7.344	3913.4E6	2540.8E6	486.266	498.107
13) T	2,4-DB	8.195	7.833	520.4E6	325.2E6	448.061	498.138
14) T	DINOSEB	9.297	8.149	1537.9E6	971.0E6	240.611	240.157
15) T	Picloram	9.123	9.090	3926.1E6	2526.8E6	341.663	334.834
16) T	DCPA	9.569	9.049	2934.0E6	1901.5E6	294.123	279.322m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS062519\
Data File : PS005207.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2019 20:30
Operator : SM\AJ
Sample : K3482-02MS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
TP-1MS

Manual Integrations
APPROVED

Ankita
6/26/2019 1:10:49 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 01:09:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

