

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070224\
 Data File : PS027114.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2024 13:47
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/03/2024
 Supervised By : Ankita Jodhani 07/03/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 03 03:55:13 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061724.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 18 02:58:29 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.212	7.752	1932.1E6	1698.7E6	759.213	733.721
Target Compounds							
1) T	Dalapon	2.622	2.699	2364.4E6	2770.9E6	668.660	625.864
2) T	3,5-DICHL...	6.386	6.704	2823.0E6	2319.8E6	701.063	690.562m
3) T	4-Nitroph...	7.011	7.279	1194.7E6	1130.7E6	676.025	680.536m
5) T	DICAMBA	7.396	7.952	7874.9E6	7468.2E6	736.406	728.476
6) T	MCPD	7.576	8.054	582.8E6	544.6E6	80.561m	74.171
7) T	MCPA	7.724	8.298	742.7E6	731.8E6	71.762	70.323
8) T	DICHLORPROP	8.094	8.665	1908.5E6	1833.1E6	693.555	693.149m
9) T	2,4-D	8.321	8.994	2127.3E6	2057.6E6	715.825m	715.693m
10) T	Pentachlo...	8.616	9.523	29482.2E6	28431.6E6	765.428m	752.773m
11) T	2,4,5-TP ...	9.191	9.897	11260.0E6	11391.5E6	754.339	742.448
12) T	2,4,5-T	9.480	10.316	11398.5E6	11608.1E6	754.866	745.229
13) T	2,4-DB	10.051	10.881	1646.1E6	1429.7E6	771.161m	719.156
14) T	DINOSEB	11.248	11.260	8260.1E6	8028.8E6	768.960	751.942
15) T	Picloram	11.061	12.354	15782.4E6	17827.6E6	750.831m	759.938
16) T	DCPA	11.544	12.302	13360.2E6	12713.3E6	704.953	685.381

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070224\
Data File : PS027114.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2024 13:47
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/03/2024
Supervised By : Ankita Jodhani 07/03/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 03:55:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061724.M
Quant Title : 8080.M
QLast Update : Tue Jun 18 02:58:29 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

