

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070319\
 Data File : PS005369.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jul 2019 09:49
 Operator : SM\AJ
 Sample : PB121021BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB121021BS

Manual Integrations
 APPROVED

Ankita
 7/8/2019 9:57:10 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 03 15:24:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.576	5.290	687.6E6	481.6E6	468.442	480.417
Target Compounds							
1) T	Dalapon	2.105	1.803	802.3E6	534.2E6	435.265m	404.989
2) T	3,5-DICHL...	4.926	4.582	896.9E6	582.3E6	446.273	441.546
3) T	4-Nitroph...	5.424	4.965	280.0E6	231.8E6	386.622	462.341
5) T	DICAMBA	5.735	5.422	2746.7E6	1773.5E6	481.481	473.773
6) T	MCPD	5.892	5.508	204.7E6	126.7E6	46.754	45.499
7) T	MCPA	6.029	5.692	275.1E6	185.3E6	43.813	45.749
8) T	DICHLORPROP	6.374	5.977	659.2E6	442.2E6	460.966	457.947
9) T	2,4-D	6.590	6.250	704.8E6	508.3E6	453.343	455.765
10) T	Pentachlo...	6.860	6.626	9649.4E6	6198.8E6	490.413	482.161
11) T	2,4,5-TP ...	7.396	6.988	3923.0E6	2530.8E6	480.871	475.251
12) T	2,4,5-T	7.670	7.350	3927.5E6	2311.5E6	488.016	453.162
13) T	2,4-DB	8.202	7.838	526.0E6	298.1E6	452.812	456.560
14) T	DINOSEB	9.302	8.152	3091.7E6	1792.9E6	483.709	443.437
15) T	Picloram	9.130	9.094	5611.5E6	3559.6E6	488.330	471.695
16) T	DCPA	9.573	9.052	4900.5E6	3194.9E6	491.262	469.325

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070319\
Data File : PS005369.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jul 2019 09:49
Operator : SM\AJ
Sample : PB121021BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleID :
PB121021BS

Manual Integrations
APPROVED

Ankita
7/8/2019 9:57:10 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 15:24:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

