

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070819\
 Data File : PS005404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2019 18:15
 Operator : SM\AJ
 Sample : K3622-02MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 OR-02-070319-AMSD

Manual Integrations
 APPROVED

Ankita
 7/9/2019 2:18:45 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 00:41:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.570	5.284	235.0E6	169.8E6	160.104	169.350
Target Compounds							
1) T	Dalapon	2.101	1.802	588.4E6	349.6E6	319.184	265.023
2) T	3,5-DICHL...	4.919	4.578	633.6E6	381.4E6	315.283	289.224
5) T	DICAMBA	5.728	5.417	2045.3E6	1327.4E6	358.528	354.594
6) T	MCPD	5.885	5.503	274.2E6	116.4E6	62.634m	41.777 #
7) T	MCPA	6.021	5.687	257.0E6	208.2E6	40.918m	51.402 #
8) T	DICHLORPROP	6.366	5.971	627.1E6	457.5E6	438.523	473.817m
9) T	2,4-D	6.581	6.243	811.7E6	609.1E6	522.128	546.161
10) T	Pentachlo...	6.852	6.620	4983.7E6	3128.8E6	253.288	243.368
11) T	2,4,5-TP ...	7.389	6.982	3847.0E6	2944.5E6	471.553	552.935
12) T	2,4,5-T	7.662	7.344	3735.8E6	2275.0E6	464.201	446.014
13) T	2,4-DB	8.195	7.833	484.0E6	289.1E6	416.715	442.874
14) T	DINOSEB	9.295	8.147	1137.8E6	671.0E6	178.019	165.946
15) T	Picloram	9.123	9.089	3730.2E6	2359.1E6	324.615	312.606m
16) T	DCPA	9.566	9.048	2868.6E6	1825.0E6	287.575	268.086m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070819\
Data File : PS005404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2019 18:15
Operator : SM\AJ
Sample : K3622-02MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
OR-02-070319-AMSD

Manual Integrations
APPROVED

Ankita
7/9/2019 2:18:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 00:41:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

