

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070919\
 Data File : PS005419.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2019 09:29
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 7/10/2019 1:27:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 12:04:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.594	5.308	1160.2E6	777.1E6	790.399	775.278
Target Compounds							
1) T	Dalapon	2.126	1.823	1294.5E6	931.6E6	702.272	706.293
2) T	3,5-DICHL...	4.944	4.601	1480.8E6	921.9E6	736.799	699.098
3) T	4-Nitroph...	5.440	4.983	484.8E6	349.7E6	669.303	697.427
5) T	DICAMBA	5.753	5.441	4468.5E6	2806.8E6	783.297	749.806
6) T	MCPD	5.914	5.531	353.8E6	213.5E6	80.821	76.637
7) T	MCPA	6.051	5.716	489.3E6	303.8E6	77.910	75.001
8) T	DICHLORPROP	6.392	5.996	1120.9E6	709.6E6	783.796	734.910
9) T	2,4-D	6.607	6.268	1170.3E6	806.2E6	752.814	722.890
10) T	Pentachlo...	6.879	6.645	14787.8E6	9238.7E6	751.558	718.616
11) T	2,4,5-TP ...	7.414	7.007	6277.1E6	3884.3E6	769.422	729.408
12) T	2,4,5-T	7.688	7.368	6292.0E6	3647.2E6	781.819	715.017
13) T	2,4-DB	8.220	7.856	866.3E6	498.2E6	745.827	763.125
14) T	DINOSEB	9.320	8.171	4913.7E6	2861.9E6	768.757	707.831
15) T	Picloram	9.148	9.113	8366.2E6	5044.7E6	728.061	668.485m
16) T	DCPA	9.591	9.071	7617.6E6	4966.7E6	763.650	729.601

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070919\
Data File : PS005419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2019 09:29
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
7/10/2019 1:27:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 12:04:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

