

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070919\
 Data File : PS005422.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2019 10:38
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 7/10/2019 1:27:08 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 12:05:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.566	5.282	1130.7E6	770.3E6	770.286m	768.506
Target Compounds							
1) T	Dalapon	2.099	1.798	1237.0E6	880.6E6	671.104m	667.562m
2) T	3,5-DICHL...	4.916	4.575	1453.1E6	924.0E6	723.057	700.706
3) T	4-Nitroph...	5.411	4.957	432.8E6	346.9E6	597.529m	691.861
5) T	DICAMBA	5.726	5.415	4401.7E6	2775.8E6	771.584	741.542
6) T	MCPD	5.886	5.505	352.0E6	211.6E6	80.419	75.980
7) T	MCPA	6.024	5.690	488.7E6	301.8E6	77.823m	74.503
8) T	DICHLORPROP	6.365	5.970	1051.2E6	705.0E6	735.051	730.154
9) T	2,4-D	6.580	6.242	1149.7E6	794.1E6	739.566	712.013
10) T	Pentachlo...	6.851	6.619	14693.0E6	9193.5E6	746.742	715.098
11) T	2,4,5-TP ...	7.387	6.981	6211.8E6	3848.3E6	761.411	722.644
12) T	2,4,5-T	7.660	7.342	6276.3E6	3616.4E6	779.878	708.982
13) T	2,4-DB	8.192	7.831	854.5E6	492.1E6	735.675	753.708
14) T	DINOSEB	9.293	8.145	4919.4E6	2831.5E6	769.649	700.312
15) T	Picloram	9.120	9.086	8038.6E6	4965.6E6	699.547	658.001
16) T	DCPA	9.564	9.045	7679.1E6	4914.8E6	769.815	721.971

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070919\
Data File : PS005422.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2019 10:38
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
7/10/2019 1:27:08 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 12:05:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

