

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071219\
 Data File : PS005481.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2019 12:53
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 7/16/2019 8:42:06 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 13:33:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.570	5.286	1098.1E6	761.2E6	748.074	759.405
Target Compounds							
1) T	Dalapon	2.102	1.802	1290.2E6	906.8E6	699.942m	687.447
2) T	3,5-DICHL...	4.920	4.579	1426.5E6	906.5E6	709.820	687.424
3) T	4-Nitroph...	5.415	4.961	440.2E6	357.0E6	607.766m	711.855
5) T	DICAMBA	5.730	5.419	4324.8E6	2738.2E6	758.108	731.500
6) T	MCPD	5.890	5.509	341.1E6	206.0E6	77.927	73.959
7) T	MCPA	6.027	5.694	474.8E6	298.4E6	75.607m	73.673
8) T	DICHLORPROP	6.369	5.974	1038.0E6	693.1E6	725.814	717.846
9) T	2,4-D	6.583	6.246	1129.2E6	783.0E6	726.358	702.078
10) T	Pentachlo...	6.855	6.623	14432.1E6	9031.0E6	733.481	702.462
11) T	2,4,5-TP ...	7.390	6.985	6058.7E6	3826.5E6	742.656	718.552
12) T	2,4,5-T	7.664	7.346	6055.5E6	3562.5E6	752.434	698.414
13) T	2,4-DB	8.196	7.835	849.2E6	484.0E6	731.099	741.330
14) T	DINOSEB	9.296	8.149	4629.7E6	2715.1E6	724.331	671.510
15) T	Picloram	9.125	9.091	7363.1E6	4586.9E6	640.762	607.821
16) T	DCPA	9.567	9.049	7326.0E6	4829.2E6	734.422	709.399

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071219\
Data File : PS005481.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2019 12:53
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
7/16/2019 8:42:06 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 13:33:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

