

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072419\
 Data File : PS005705.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2019 18:36
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 7/25/2019 4:18:18 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 25 01:00:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071819.M
 Quant Title : 8080.M
 QLast Update : Fri Jul 19 08:54:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.579	5.297	1072.1E6	796.9E6	680.381	771.529
Target Compounds							
1) T	Dalapon	2.112	1.810	1361.1E6	987.1E6	696.381m	684.448
2) T	3,5-DICHL...	4.931	4.591	1363.0E6	920.2E6	640.631m	678.858
3) T	4-Nitroph...	5.425	4.973	426.8E6	357.9E6	597.026m	722.251
5) T	DICAMBA	5.738	5.429	4165.5E6	2873.0E6	673.450m	726.039
6) T	MCPD	5.898	5.519	327.9E6	218.4E6	67.504m	75.370
7) T	MCPA	6.036	5.704	454.8E6	315.6E6	67.278m	71.636
8) T	DICHLORPROP	6.377	5.983	971.1E6	720.6E6	638.897m	706.452
9) T	2,4-D	6.591	6.255	1097.5E6	820.5E6	625.798	702.669
10) T	Pentachlo...	6.862	6.631	14292.9E6	9498.3E6	660.117	706.157
11) T	2,4,5-TP ...	7.398	6.994	5774.2E6	3974.8E6	652.543	713.169
12) T	2,4,5-T	7.672	7.355	5738.2E6	3715.1E6	677.812	723.987
13) T	2,4-DB	8.203	7.843	785.6E6	505.4E6	627.100	733.621
14) T	DINOSEB	9.302	8.158	5041.4E6	2968.2E6	722.740	733.140
15) T	Picloram	9.131	9.099	7287.8E6	4803.9E6	584.961	613.301
16) T	DCPA	9.574	9.057	7701.4E6	5059.0E6	707.248	715.708

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072419\
Data File : PS005705.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2019 18:36
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
7/25/2019 4:18:18 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 01:00:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071819.M
Quant Title : 8080.M
QLast Update : Fri Jul 19 08:54:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

