

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072823\
 Data File : PS023852.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2023 08:20
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 22:56:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
 Quant Title : 8080.M
 QLast Update : Thu Jul 20 19:21:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.371	6.806	1478.5E6	500.3E6	745.729	735.627
Target Compounds							
1) T	Dalapon	2.489	2.560	1436.4E6	730.3E6	670.600	659.470
2) T	3,5-DICHL...	5.744	5.985	2147.6E6	733.1E6	695.438	684.570
3) T	4-Nitroph...	6.208	6.417	886.1E6	288.1E6	712.771	680.148
5) T	DICAMBA	6.513	6.970	6506.6E6	2176.7E6	718.206	713.252
8) T	DICHLORPROP	7.103	7.587	1642.2E6	522.6E6	702.912	691.793
9) T	2,4-D	7.300	7.873	1819.3E6	591.2E6	705.196	693.591
10) T	Pentachlo...	7.531	8.299	23518.8E6	7981.2E6	731.414	721.562
11) T	2,4,5-TP ...	8.053	8.658	9361.7E6	3009.2E6	723.931	724.218
12) T	2,4,5-T	8.308	9.030	8746.9E6	2868.9E6	742.728	728.753
13) T	2,4-DB	8.815	9.534	1116.0E6	374.0E6	736.306	754.193
14) T	DINOSEB	9.871	9.863	6343.1E6	2098.3E6	767.651	759.025
15) T	Picloram	9.694	10.819	11349.5E6	4162.3E6	740.540	686.261
16) T	DCPA	10.135	10.787	10502.4E6	3460.1E6	719.635	727.751

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072823\
Data File : PS023852.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2023 08:20
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 22:56:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
Quant Title : 8080.M
QLast Update : Thu Jul 20 19:21:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

