

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082119\
 Data File : PS006274.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2019 14:20
 Operator : SM\AJ
 Sample : K4404-06MSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 T1-L-(0-2.5)MSD

Manual Integrations
 APPROVED

Ankita
 8/22/2019 2:46:33 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 01:50:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.485	5.214	496.2E6	284.7E6	574.887	578.848
Target Compounds							
1) T	Dalapon	2.045	1.756	423.5E6	375.6E6	388.686	598.596 #
2) T	3,5-DICHL...	4.850	4.516	616.6E6	349.7E6	571.876	577.370
3) T	4-Nitroph...	5.341	4.900	205.1E6	111.1E6	546.376	448.436
5) T	DICAMBA	5.636	5.342	1818.2E6	1032.2E6	566.118	559.590
6) T	MCPD	5.785	5.424	134.5E6	74846832	59.748m	60.502
7) T	MCPA	5.919	5.603	200.1E6	129.9E6	63.383	63.538
8) T	DICHLORPROP	6.266	5.883	487.4E6	284.2E6	642.724	639.324
9) T	2,4-D	6.482	6.154	606.3E6	385.9E6	744.078	701.961
10) T	Pentachlo...	6.741	6.520	7304.4E6	4138.8E6	617.927	612.262
11) T	2,4,5-TP ...	7.276	6.884	2988.1E6	1620.2E6	655.030	622.664
12) T	2,4,5-T	7.551	7.243	3315.2E6	1575.8E6	670.470	633.837m
13) T	2,4-DB	8.080	7.730	475.2E6	206.2E6	763.076	619.454
14) T	DINOSEB	9.168	8.039	2359.2E6	1165.8E6	644.443	606.901
15) T	Picloram	9.004	8.978	4314.0E6	2552.4E6	579.796	565.832
16) T	DCPA	9.441	8.934	3932.4E6	2205.5E6	678.511	632.479m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082119\
Data File : PS006274.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2019 14:20
Operator : SM\AJ
Sample : K4404-06MSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
T1-L-(0-2.5)MSD

Manual Integrations
APPROVED

Ankita
8/22/2019 2:46:33 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 01:50:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

