

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082119\
 Data File : PS006276.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2019 15:04
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 01:51:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.490	5.220	683.2E6	384.4E6	791.432	781.482
Target Compounds							
1) T	Dalapon	2.054	1.764	706.4E6	410.1E6	648.291	653.533
2) T	3,5-DICHL...	4.857	4.522	804.6E6	449.9E6	746.285	742.746
3) T	4-Nitroph...	5.346	4.906	278.4E6	184.0E6	741.709	742.913
5) T	DICAMBA	5.642	5.348	2389.7E6	1352.6E6	744.070	733.291
6) T	MCPD	5.792	5.431	164.7E6	96065344	73.183	77.653
7) T	MCPA	5.925	5.610	243.1E6	153.9E6	76.988	75.294
8) T	DICHLORPROP	6.272	5.889	562.3E6	330.8E6	741.474	744.020
9) T	2,4-D	6.490	6.161	591.4E6	405.2E6	725.804	737.127
10) T	Pentachlo...	6.746	6.527	8999.7E6	5063.2E6	761.339	748.998
11) T	2,4,5-TP ...	7.283	6.890	3447.3E6	1953.5E6	755.684	750.726
12) T	2,4,5-T	7.557	7.250	3916.7E6	1806.6E6	792.115	726.710
13) T	2,4-DB	8.088	7.736	472.7E6	247.9E6	759.002	745.018
14) T	DINOSEB	9.174	8.045	2833.3E6	1423.3E6	773.966	740.992
15) T	Picloram	9.009	8.984	5537.5E6	3183.7E6	744.228	705.784
16) T	DCPA	9.446	8.940	4533.7E6	2674.4E6	782.270	766.940

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082119\
Data File : PS006276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2019 15:04
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 01:51:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

