

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082923\
 Data File : PS024192.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Aug 2023 09:02
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 19:20:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 28 16:48:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.364	6.792	1520.3E6	514.6E6	763.085	766.275
Target Compounds							
1) T	Dalapon	2.481	2.548	1485.6E6	717.5E6	683.518	680.369
2) T	3,5-DICHL...	5.738	5.975	2160.8E6	751.7E6	710.828	720.897
3) T	4-Nitroph...	6.201	6.405	876.5E6	293.0E6	705.535	694.711
5) T	DICAMBA	6.506	6.955	6356.0E6	2112.0E6	729.135	733.292
6) T	MCPD	6.658	7.052	470.8E6	142.1E6	75.603	74.154
7) T	MCPA	6.781	7.259	637.3E6	213.6E6	73.427	73.087
8) T	DICHLORPROP	7.094	7.570	1632.9E6	524.1E6	718.285	721.309
9) T	2,4-D	7.291	7.854	1906.2E6	641.9E6	716.655	720.520
10) T	Pentachlo...	7.521	8.280	23417.1E6	7857.9E6	748.805	741.684
11) T	2,4,5-TP ...	8.043	8.639	9851.0E6	3063.7E6	736.979	738.388
12) T	2,4,5-T	8.297	9.009	9159.4E6	2965.7E6	739.256	739.015
13) T	2,4-DB	8.802	9.513	1567.5E6	367.0E6	727.171	798.437
14) T	DINOSEB	9.857	9.841	6943.7E6	2156.8E6	718.626	721.490
15) T	Picloram	9.681	10.794	11894.1E6	4252.6E6	728.308	716.070
16) T	DCPA	10.122	10.764	12010.5E6	3837.1E6	741.669	747.980

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS082923\
Data File : PS024192.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Aug 2023 09:02
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 19:20:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
Quant Title : 8080.M
QLast Update : Mon Aug 28 16:48:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

